



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119651

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/13/2022	08:20
2	CAL2	1	Water	NA	NA	20.3	7.00	04/13/2022	08:21
3	CAL3	1	Water	NA	NA	20.3	10.03	04/13/2022	08:22
4	ICV	1	Water	NA	NA	20.4	6.98	04/13/2022	08:25
5	CCV1	1	Water	NA	NA	20.3	2.01	04/13/2022	08:26
6	N2378-01	1	Water	NA	NA	20.1	6.80	04/13/2022	08:29
7	N2378-01DUP	1	Water	NA	NA	20.2	6.81	04/13/2022	08:30
8	CCV2	1	Water	NA	NA	20.4	12.03	04/13/2022	08:31

1B 119651

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2378 WorkList ID : 158630 Department : Wet-Chemistry Date : 04-13-2022 08:13:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/12/2022	Water	N2378-01	pH		Cool 4 deg C	PSEG03 O11	T-2432	04/12/2022	9040C

Date/Time 04/13/22 08:13
Raw Sample Received by: JP/WR
Raw Sample Relinquished by: CR SM

Date/Time 04/13/22 09:00
Raw Sample Received by: CR SM
Raw Sample Relinquished by: JP (WC)