

LB119652


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

59071



A= 0.0000e+000

B= 1.6934e-004

C= -2.0977e-002

Rho= 0.9999817

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	-0.011	-0.011	61	0.000	61					
0.2	0.200	0.225	0.025	1452	0.0 %	1452					13
2.5	2.500	2.474	-0.026	14734	0.0 %	14734					-1
5.0	5.000	4.994	-0.006	29617	0.0 %	29617					0
7.5	7.500	7.535	0.035	44621	0.0 %	44621					0
10.0	10.000	9.982	-0.018	59071	0.0 %	59071					0

LB119652 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 13 Apr 2022 11:05:36			
Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type
0.0 - 1	50	61	-	0.00007471B	PPB	13 Apr 2022 11:11:31	S Std
0.2 - 1	50-2	1452	-	0.20007471B	PPB	13 Apr 2022 11:13:48	S Std
2.5 - 1	52.5	14734	-	2.50007471B	PPB	13 Apr 2022 11:16:04	S Std
5.0 - 1	55	29617	-	5.00007471B	PPB	13 Apr 2022 11:18:21	S Std
7.5 - 1	57.5	44621	-	7.50007471B	PPB	13 Apr 2022 11:20:38	S Std
10.0 - 1	50	59071	-	10.00007471B	PPB	13 Apr 2022 11:22:54	S Std
ICV76 - 1	ICV76	22436	3.7783	-7471B	PPB	13 Apr 2022 11:25:40	U SMPL
ICB76 - 1	ICB76	-49	-0.0293	-7471B	PPB	13 Apr 2022 11:27:54	U SMPL
CCV46 - 1	CCV46	29448	4.9658	-7471B	PPB	13 Apr 2022 11:35:19	U SMPL
CCB46 - 1	CCB46	-56	-0.0305	-7471B	PPB	13 Apr 2022 11:37:34	U SMPL
CRA - 1	CRA	1414	0.2185	-7471B	PPB	13 Apr 2022 11:39:49	U SMPL
HighStd - 1	HighStd	58758	9.9291	-7471B	PPB	13 Apr 2022 11:42:04	U SMPL
ChkStd - 1	ChkStd	42006	7.0923	-7471B	PPB	13 Apr 2022 11:44:19	U SMPL
PB144090BL - 1	PBS	-129	-0.0428	-7471B	PPB	13 Apr 2022 11:46:36	U SMPL
PB144090BS - 1	LCSS	22504	3.7899	-7471B	PPB	13 Apr 2022 11:51:46	U SMPL
N2349-13 - 1	DEL-7	1527	0.2376	-7471B	PPB	13 Apr 2022 11:54:01	U SMPL
N2349-15 - 1	DEL-8	5253	0.8686	-7471B	PPB	13 Apr 2022 11:56:18	U SMPL
N2349-17 - 1	DEL-8D	10491	1.7556	-7471B	PPB	13 Apr 2022 11:58:35	U SMPL
N2349-19 - 1	COM-1	2876	0.4660	-7471B	PPB	13 Apr 2022 12:00:51	U SMPL
N2360-01 - 1	JC-03-041122-A	5098	0.8423	-7471B	PPB	13 Apr 2022 12:03:06	U SMPL
CCV47 - 1	CCV47	29313	4.9429	-7471B	PPB	13 Apr 2022 12:05:21	U SMPL
CCB47 - 1	CCB47	-41	-0.0279	-7471B	PPB	13 Apr 2022 12:07:39	U SMPL
N2360-01DUP - 1	JC-03-041122-ADUP	4648	0.7661	-7471B	PPB	13 Apr 2022 12:15:02	U SMPL
N2360-01MS - 1	JC-03-041122-AMS	27597	4.6523	-7471B	PPB	13 Apr 2022 12:19:51	U SMPL
N2360-01MSD - 1	JC-03-041122-AMSD	27924	4.7077	-7471B	PPB	13 Apr 2022 12:22:06	U SMPL
N2360-03 - 1	JC-03-041122-B	4509	0.7426	-7471B	PPB	13 Apr 2022 12:24:22	U SMPL
N2362-01 - 1	WC-14A-1A-4-1-1	35557	6.0003	-7471B	PPB	13 Apr 2022 12:26:38	U SMPL
N2362-02 - 1	WC-14A-1A-4-1-2	59199	10.0038	-7471B	PPB	13 Apr 2022 12:28:54	U SMPL
N2362-03 - 1	WC-14A-1A-4-1-3	1292	0.1978	-7471B	PPB	13 Apr 2022 12:31:11	U SMPL
N2362-04 - 1	WC-14A-1A-4-1-4	1668	0.2615	-7471B	PPB	13 Apr 2022 12:33:29	U SMPL
N2365-01 - 1	TR-04-041122	2708	0.4376	-7471B	PPB	13 Apr 2022 12:35:46	U SMPL
N2367-01 - 1	AU-05-041122	1796	0.2832	-7471B	PPB	13 Apr 2022 12:38:01	U SMPL
CCV48 - 1	CCV48	30132	5.0816	-7471B	PPB	13 Apr 2022 12:40:16	U SMPL
CCB48 - 1	CCB48	-31	-0.0262	-7471B	PPB	13 Apr 2022 12:42:32	U SMPL
N2368-02 - 1	E14ST-EB3-041122-0-5	11363	1.9032	-7471B	PPB	13 Apr 2022 12:44:47	U SMPL
N2368-03 - 1	E14ST-EB3-041122-5-10	5093	0.8415	-7471B	PPB	13 Apr 2022 12:47:02	U SMPL
N2369-01 - 1	EMERSON-SCHOOL-STOCK-PILE	4035	0.6623	-7471B	PPB	13 Apr 2022 12:49:18	U SMPL
N2376-01 - 1	NB-07-041222	4401	0.7243	-7471B	PPB	13 Apr 2022 12:51:33	U SMPL
N2380-01 - 1	OK-02-041222	2040	0.3245	-7471B	PPB	13 Apr 2022 12:53:49	U SMPL
N2381-01 - 1	26572	5463	0.9041	-7471B	PPB	13 Apr 2022 12:56:05	U SMPL
N2360-01LX5 - 1		1005	0.1492	-7471B	PPB	13 Apr 2022 12:58:21	U SMPL
N2360-01A - 1		28808	4.8574	-7471B	PPB	13 Apr 2022 13:07:38	U SMPL
N2362-02DLX2 - 1		29158	4.9166	-7471B	PPB	13 Apr 2022 13:09:55	U SMPL
CCV49 - 1	CCV49	29767	5.0198	-7471B	PPB	13 Apr 2022 13:12:11	U SMPL
CCB49 - 1	CCB49	-74	-0.0335	-7471B	PPB	13 Apr 2022 13:14:28	U SMPL