



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119654

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	27.00	80.6	80.9	04/13/2022	11:00
2	N2340-11		1	9	100.00	>212.0	>212.0	04/13/2022	11:33
3	N2378-01		1	10	100.00	>212.0	>212.0	04/13/2022	11:49
4	N2385-02		1	10	100.00	>212.0	>212.0	04/13/2022	12:05
5	N2385-02DUP		1	11	100.00	>212.0	>212.0	04/13/2022	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 041322 WorkList ID : 158649 Department : Wet-Chemistry Date : 04-13-2022 10:16:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/15/2022	Water	N2340-11	Flash Point	Cool 4 deg C	LIRO01	N13	TWP-08	04/07/2022	1010B
04/13/2022	Water	N2378-01	Flash Point	Cool 4 deg C	PSEG03	O11	T-2432	04/12/2022	1010B
04/14/2022	Water	N2385-02	Flash Point	Cool 4 deg C	PSEG03	O11	N48963	04/13/2022	1010B

Date/Time 4/13/22 11:15
 Raw Sample Received by: AP (WOC)
 Raw Sample Relinquished by: AP (WOC)

Date/Time 4/13/22 12:40
 Raw Sample Received by: AP (WOC)
 Raw Sample Relinquished by: AP (WOC)