



## Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119661

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2743     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 9                 | 28.00      | 82.4      | 82.7            | 04/13/2022 | 14:10     |
| 2   | N2009-01    |               | 1  | 5                 | 17.00      | 62.6      | 62.9            | 04/13/2022 | 14:20     |
| 3   | N2009-01DUP |               | 1  | 6                 | 18.00      | 64.4      | 64.7            | 04/13/2022 | 14:33     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 041322-2      WorkList ID : 158656      Department : Wet-Chemistry      Date : 04-13-2022 14:18:07

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 03/24/2022 | Solid  | N2009-01 | Flash Point | Cool 4 deg C | PSEG03   | E11                         | 13563           | 03/17/2022   | 1010A  |

Date/Time 4/13/22 14:20  
 Raw Sample Received by: CP (WC)  
 Raw Sample Relinquished by: CP SP

Date/Time 4/13/22 14:45  
 Raw Sample Received by: CP SP  
 Raw Sample Relinquished by: CP (WC)