

LB119663


7470A

INSTRUMENT NO: CV2

Linear

 μ Abs.:

42789



A= 0.0000e+000

B= 2.3054e-004

C= -2.6482e-002

Rho= 0.9994830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.003	0.003	127	0.000	127					-
0.2	0.200	0.188	-0.012	932	0.0 %	932					-6
2.5	2.500	2.377	-0.123	10426	0.0 %	10426					-5
5.0	5.000	5.143	0.143	22424	0.0 %	22424					3
7.5	7.500	7.650	0.150	33298	0.0 %	33298					2
10.0	10.000	9.838	-0.162	42789	0.0 %	42789					-2

LB119663 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Apr 2022 14:14:55					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	56	127	-	0.0000	7470A	PPB	13 Apr 2022 14:30:30	S	Std
0.2 - 1	57.2	932	-	0.2000	7470A	PPB	13 Apr 2022 14:32:47	S	Std
2.5 - 1	52.5	10426	-	2.5000	7470A	PPB	13 Apr 2022 14:35:04	S	Std
5.0 - 1	55	22424	-	5.0000	7470A	PPB	13 Apr 2022 14:37:20	S	Std
7.5 - 1	57.5	33298	-	7.5000	7470A	PPB	13 Apr 2022 14:39:37	S	Std
10.0 - 1	51.0	42789	-	10.0000	7470A	PPB	13 Apr 2022 14:41:52	S	Std
ICV77 - 1	ICV77	16339	3.7404	-	7470A	PPB	13 Apr 2022 14:44:38	U	SMPL
ICB77 - 1	ICB77	-85	-0.0461	-	7470A	PPB	13 Apr 2022 14:46:53	U	SMPL
CCV50 - 1	CCV50	22218	5.0957	-	7470A	PPB	13 Apr 2022 14:49:08	U	SMPL
CCB50 - 1	CCB50	-87	-0.0465	-	7470A	PPB	13 Apr 2022 14:51:23	U	SMPL
CRA - 1	CRA	949	0.1923	-	7470A	PPB	13 Apr 2022 14:53:38	U	SMPL
HighStd - 1	HighStd	43274	9.9501	-	7470A	PPB	13 Apr 2022 14:55:53	U	SMPL
ChkStd - 1	ChkStd	31331	7.1967	-	7470A	PPB	13 Apr 2022 14:58:08	U	SMPL
PB144094BL - 1	PBW	203	0.0203	-	7470A	PPB	13 Apr 2022 15:02:29	U	SMPL
PB144094BS - 1	LCSW	16426	3.7604	-	7470A	PPB	13 Apr 2022 15:04:44	U	SMPL
N2321-02 - 1	RW7A-GW-20220406	459	0.0793	-	7470A	PPB	13 Apr 2022 15:07:00	U	SMPL
N2368-05 - 1	FIELD-BLANK	286	0.0395	-	7470A	PPB	13 Apr 2022 15:09:17	U	SMPL
N2378-01 - 1	T-2432	844	0.1681	-	7470A	PPB	13 Apr 2022 15:11:33	U	SMPL
N2378-01DUP - 1	T-2432DUP	842	0.1676	-	7470A	PPB	13 Apr 2022 15:16:19	U	SMPL
N2378-01MS - 1	T-2432MS	16470	3.7706	-	7470A	PPB	13 Apr 2022 15:21:05	U	SMPL
CCV51 - 1	CCV51	22379	5.1329	-	7470A	PPB	13 Apr 2022 15:23:21	U	SMPL
CCB51 - 1	CCB51	-92	-0.0477	-	7470A	PPB	13 Apr 2022 15:25:36	U	SMPL
N2378-01MSD - 1	T-2432MSD	16488	3.7747	-	7470A	PPB	13 Apr 2022 15:27:51	U	SMPL
N2385-02 - 1	N48963	187	0.0166	-	7470A	PPB	13 Apr 2022 15:30:06	U	SMPL
N2378-01LX5 - 1		171	0.0129	-	7470A	PPB	13 Apr 2022 15:32:22	U	SMPL
N2378-01A - 1		17323	3.9672	-	7470A	PPB	13 Apr 2022 15:34:37	U	SMPL
CCV52 - 1	CCV52	22210	5.0939	-	7470A	PPB	13 Apr 2022 15:36:53	U	SMPL
CCB52 - 1	CCB52	-79	-0.0447	-	7470A	PPB	13 Apr 2022 15:39:10	U	SMPL