



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119667

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	04/13/2022	15:35
2	CAL2	1	Water	NA	NA	20.4	7.00	04/13/2022	15:36
3	CAL3	1	Water	NA	NA	20.4	10.03	04/13/2022	15:37
4	ICV	1	Water	NA	NA	20.3	7.01	04/13/2022	15:39
5	CCV1	1	Water	NA	NA	20.4	2.01	04/13/2022	15:40
6	N2385-02	1	Water	NA	NA	20.1	7.40	04/13/2022	15:42
7	N2385-02DUP	1	Water	NA	NA	20.2	7.41	04/13/2022	15:43
8	N2389-01	1	Water	NA	NA	21.1	7.58	04/13/2022	15:46
9	N2389-03	1	Water	NA	NA	20.7	6.86	04/13/2022	15:50
10	N2389-05	1	Water	NA	NA	21.0	5.63	04/13/2022	15:51
11	N2389-06	1	Water	NA	NA	21.0	5.78	04/13/2022	15:53
12	CCV2	1	Water	NA	NA	20.5	12.03	04/13/2022	16:00

W 119667

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2389 WorkList ID : 158660 Department : Wet-Chemistry Date : 04-13-2022 15:24:17

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
04/13/2022	Water	N2385-02	pH <i>A</i>	Cool 4 deg C	PSEG03	O11	N48963	04/13/2022	9040C
04/13/2022	Water	N2389-01	pH <i>A</i>	Cool 4 deg C	PSEG03	O11	20564	04/13/2022	9040C
04/13/2022	Water	N2389-03	pH <i>A</i>	Cool 4 deg C	PSEG03	O11	33630	04/13/2022	9040C
04/13/2022	Water	N2389-05	pH	Cool 4 deg C	PSEG03	O11	33637	04/13/2022	9040C
04/13/2022	Water	N2389-06	pH	Cool 4 deg C	PSEG03	O11	33638	04/13/2022	9040C

Date/Time 04.13.22 15:30
Raw Sample Received by: DP
Raw Sample Relinquished by: CP SM

Date/Time 04.13.22 16:35
Raw Sample Received by: CP SM
Raw Sample Relinquished by: DP (acc)