



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119669

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/13/2022	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/13/2022	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	04/13/2022	16:17
4	ICV	1	Water	NA	NA	20.5	6.98	04/13/2022	16:20
5	CCV1	1	Water	NA	NA	20.7	2.01	04/13/2022	16:22
6	N2389-03	1	Water	NA	NA	20.7	6.86	04/13/2022	16:23
7	N2389-03DUP	1	Water	NA	NA	20.8	6.87	04/13/2022	16:24
8	N2396-02	1	Water	NA	NA	20.4	8.42	04/13/2022	16:27
9	N2396-03	1	Water	NA	NA	20.6	7.51	04/13/2022	16:29
10	CCV2	1	Water	NA	NA	20.4	12.03	04/13/2022	16:33

19119669

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N2389

WorkList ID : 158662

Department : Wet-Chemistry

Date : 04-13-2022 16:07:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/13/2022	Water	N2389-03	Corrosivity	Cool 4 deg C	PSEG03	O11	33630	04/13/2022	9040C
04/13/2022	Water	N2396-02	Corrosivity	Cool 4 deg C	PSEG03	O11	5576	04/13/2022	9040C
04/13/2022	Water	N2396-03	Corrosivity	Cool 4 deg C	PSEG03	O11	5581	04/13/2022	9040C

16:10

Date/Time 04.13.22
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: SP (WC)

191100

Date/Time 04.13.22
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: SP (WC)