



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119670

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	04/13/2022	15:30
2	CAL2	1	Water	NA	NA	20.4	6.98	04/13/2022	15:31
3	CAL3	1	Water	NA	NA	20.4	10.03	04/13/2022	15:32
4	ICV	1	Water	NA	NA	20.5	7.01	04/13/2022	15:35
5	CCV1	1	Water	NA	NA	20.4	2.01	04/13/2022	15:36
6	N2364-02	1	Solid	20.03	20	20.4	8.95	04/13/2022	15:39
7	N2364-02DUP	1	Solid	20.04	20	20.5	8.96	04/13/2022	15:40
8	N2364-04	1	Solid	20.01	20	20.5	8.08	04/13/2022	15:42
9	N2364-06	1	Solid	20.03	20	20.6	7.13	04/13/2022	15:44
10	N2364-08	1	Solid	20.05	20	20.4	9.21	04/13/2022	15:45
11	N2392-02	1	Solid	20.03	20	21.4	9.60	04/13/2022	15:47
12	N2396-01	1	Solid	20.01	20	20.2	5.76	04/13/2022	15:48
13	CCV2	1	Water	NA	NA	20.5	12.02	04/13/2022	15:50

WB 119670

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2392 WorkList ID : 158659 Department : Wet-Chemistry Date : 04-13-2022 15:23:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/11/2022	Solid	N2364-02	Corrosivity	Cool 4 deg C	LIRO01	O11	SB-01-COMP	04/11/2022	9045D
04/11/2022	Solid	N2364-04	Corrosivity	Cool 4 deg C	LIRO01	O11	SB-02-COMP	04/11/2022	9045D
04/11/2022	Solid	N2364-06	Corrosivity	Cool 4 deg C	LIRO01	O11	SB-03-COMP	04/11/2022	9045D
04/11/2022	Solid	N2364-08	Corrosivity	Cool 4 deg C	LIRO01	O11	SB-09-COMP	04/11/2022	9045D
04/13/2022	Solid	N2392-02	Corrosivity	Cool 4 deg C	PSEG03	O11	VNJ-212	04/13/2022	9045D
04/13/2022	Solid	N2396-01	Corrosivity	Cool 4 deg C	PSEG03	O11	5575	04/13/2022	9045D

Date/Time 04-13-22 15:25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 04-13-22 17:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]