


LB119674

245.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5269e-004

C= -1.2997e-002

Rho= 0.9997187

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.025	-0.025	-49	0.000	-49					-2
0.05	0.050	0.049	-0.001	246	0.0 %	246	0				18
0.2	0.200	0.236	0.036	985	0.0 %	985					-4
2.5	2.500	2.404	-0.096	9566	0.0 %	9566					-4
5.0	5.000	5.184	0.184	20568	0.0 %	20568					-1
7.5	7.500	7.404	-0.096	29353	0.0 %	29353					0
10.0	10.000	9.997	-0.003	39615	0.0 %	39615					

LB119674 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 14 Apr 2022 12:05:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-49	-	0.0000	245.1	PPB	14 Apr 2022 13:44:18	S	Std
0.05 - 1	50.05	246	-	0.0500	245.1	PPB	14 Apr 2022 13:46:35	S	Std
0.2 - 1	51.2	985	-	0.2000	245.1	PPB	14 Apr 2022 13:48:51	S	Std
2.5 - 1	52.5	9566	-	2.5000	245.1	PPB	14 Apr 2022 13:51:09	S	Std
5.0 - 1	55	20568	-	5.0000	245.1	PPB	14 Apr 2022 13:53:25	S	Std
7.5 - 1	57.5	29353	-	7.5000	245.1	PPB	14 Apr 2022 13:55:41	S	Std
10.0 - 1	510	39615	-	10.0000	245.1	PPB	14 Apr 2022 13:57:57	S	Std
ICV78 - 1	ICV78	15429	3.8858	-245.1	PPB	PPB	14 Apr 2022 14:06:25	U	SMPL
ICB78 - 1	ICB78	-52	-0.0261	-245.1	PPB	PPB	14 Apr 2022 14:08:39	U	SMPL
CCV53 - 1	CCV53	20627	5.1993	-245.1	PPB	PPB	14 Apr 2022 14:10:54	U	SMPL
CCB53 - 1	CCB53	-45	-0.0244	-245.1	PPB	PPB	14 Apr 2022 14:13:09	U	SMPL
CRA - 1	CRA	269	0.0550	-245.1	PPB	PPB	14 Apr 2022 14:15:25	U	SMPL
HighStd - 1	HighStd	37827	9.5456	-245.1	PPB	PPB	14 Apr 2022 14:20:20	U	SMPL
ChkStd - 1	ChkStd	28024	7.0684	-245.1	PPB	PPB	14 Apr 2022 14:25:12	U	SMPL
PB144118BL - 1	PBW	-137	-0.0476	-245.1	PPB	PPB	14 Apr 2022 14:27:27	U	SMPL
PB144118BS - 1	LCSW	14844	3.7380	-245.1	PPB	PPB	14 Apr 2022 14:32:25	U	SMPL
N2284-04 - 1	HP-627	78	0.0067	-245.1	PPB	PPB	14 Apr 2022 14:34:41	U	SMPL
N2284-04DUP - 1	HP-627DUP	112	0.0153	-245.1	PPB	PPB	14 Apr 2022 14:36:57	U	SMPL
N2284-04MS - 1	HP-627MS	17339	4.3684	-245.1	PPB	PPB	14 Apr 2022 14:39:13	U	SMPL
N2284-04MSD - 1	HP-627MSD	17415	4.3876	-245.1	PPB	PPB	14 Apr 2022 14:41:30	U	SMPL
N2294-01 - 1	402	46	-0.0014	-245.1	PPB	PPB	14 Apr 2022 14:43:45	U	SMPL
CCV54 - 1	CCV54	20874	5.2617	-245.1	PPB	PPB	14 Apr 2022 14:46:00	U	SMPL
CCB54 - 1	CCB54	-59	-0.0279	-245.1	PPB	PPB	14 Apr 2022 14:48:15	U	SMPL
N2340-11 - 1	TWP-08	145	0.0236	-245.1	PPB	PPB	14 Apr 2022 14:50:30	U	SMPL
N2284-04LX5 - 1		47	-0.0011	-245.1	PPB	PPB	14 Apr 2022 14:52:47	U	SMPL
N2284-04A - 1		17532	4.4172	-245.1	PPB	PPB	14 Apr 2022 14:55:02	U	SMPL
CCV55 - 1	CCV55	20724	5.2238	-245.1	PPB	PPB	14 Apr 2022 14:57:18	U	SMPL
CCB55 - 1	CCB55	-57	-0.0274	-245.1	PPB	PPB	14 Apr 2022 14:59:34	U	SMPL