



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119680

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/14/2022	12:20
2	CAL2	1	Water	NA	NA	20.7	7.01	04/14/2022	12:21
3	CAL3	1	Water	NA	NA	20.7	10.03	04/14/2022	12:22
4	ICV	1	Water	NA	NA	20.6	6.99	04/14/2022	12:24
5	CCV1	1	Water	NA	NA	20.7	2.01	04/14/2022	12:25
6	N2420-02	1	Water	NA	NA	20.2	5.01	04/14/2022	12:30
7	N2423-07	1	Water	NA	NA	22.4	6.68	04/14/2022	12:33
8	N2423-07DUP	1	Water	NA	NA	22.5	6.69	04/14/2022	12:34
9	CCV2	1	Water	NA	NA	20.9	12.03	04/14/2022	12:35

VB 119680

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2423 WorkList ID : 158698 Department : Wet-Chemistry Date : 04-14-2022 13:03:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
04/13/2022	Water	N2420-02	pH <u>E</u>	Cool 4 deg C	TETR06	N21	RW4-IDWGW-20220413	04/13/2022	9040C
04/14/2022	Water	N2423-07	pH <u>D</u>	Cool 4 deg C	PSEG03	O11	D2456-D2459	04/14/2022	9040C

Date/Time 04.14.22 12:15
Raw Sample Received by: PLC
Raw Sample Relinquished by: JR(SM)

Date/Time 04.14.22 13:00
Raw Sample Received by: JR(SM)
Raw Sample Relinquished by: JR(COC)