



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119682

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP96281
calibration std. phosphate 0.5 ppm	WP96280
calibration std. phosphate 0.3 ppm	WP96279
calibration std. phosphate 0.1 ppm	WP96278
calibration std. phosphate 0.05 ppm	WP96277
calibration std. 0 ppm	WP96276
phosphate CCV std.	WP96282
Combined reagent	WP96321
Phenolphthalein indicator	WP93767
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP96283

Intercept: -0.0009

Slope: 0.6495

Regression: 0.999964

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.001		04/14/2022	13:35
2	CAL2	0.05	1	50	50	0.033	0.052	-4	04/14/2022	13:35
3	CAL3	0.10	1	50	50	0.065	0.101	-1	04/14/2022	13:36
4	CAL4	0.30	1	50	50	0.190	0.294	2	04/14/2022	13:36
5	CAL5	0.50	1	50	50	0.323	0.499	0.2	04/14/2022	13:37
6	CAL6	1.00	1	50	50	0.650	1.002	-0.2	04/14/2022	13:37



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119682

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.496	04/14/2022	13:38
2	ICB		1	50	50	0.001	0.003	04/14/2022	13:38
3	CCV1	0.50	1	50	50	0.324	0.500	04/14/2022	13:39
4	CCB1		1	50	50	0.002	0.004	04/14/2022	13:39
5	PB144107BL		1	50	50	0.002	0.004	04/14/2022	13:40
6	PB144107BS	0.50	1	50	50	0.330	0.509	04/14/2022	13:40
7	N2363-01		1	50	50	0.916	1.412	04/14/2022	13:41
8	N2397-01		1	50	50	0.661	1.019	04/14/2022	13:41
9	N2397-01DUP		1	50	50	0.660	1.018	04/14/2022	13:42
10	N2397-01MS	0.50	1	50	50	0.724	1.116	04/14/2022	13:42
11	N2397-01MSD	0.50	1	50	50	0.726	1.119	04/14/2022	13:43
12	N2400-01		1	50	50	1.034	1.593	04/14/2022	13:43
13	N2363-01		10	50	50	0.198	0.306	04/14/2022	13:44
14	N2397-01		5	50	50	0.143	0.222	04/14/2022	13:44
15	CCV2	0.50	1	50	50	0.321	0.496	04/14/2022	13:45
16	CCB2		1	50	50	0.003	0.006	04/14/2022	13:45
17	N2397-01DUP		5	50	50	0.142	0.220	04/14/2022	13:46
18	N2397-01MS	0.50	5	50	50	0.215	0.332	04/14/2022	13:46
19	N2397-01MSD	0.50	5	50	50	0.213	0.329	04/14/2022	13:47
20	N2400-01		10	50	50	0.232	0.359	04/14/2022	13:47
21	CCV3	0.50	1	50	50	0.326	0.503	04/14/2022	13:48
22	CCB3		1	50	50	0.001	0.003	04/14/2022	13:48