



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119688

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	04/14/2022	14:10
2	N2389-01		1	10	100.00	>212.0	>212.0	04/14/2022	14:30
3	N2389-03		1	9	100.00	>212.0	>212.0	04/14/2022	14:36
4	N2389-07		1	10	100.00	>212.0	>212.0	04/14/2022	14:51
5	N2396-02		1	10	100.00	>212.0	>212.0	04/14/2022	15:10
6	N2396-03		1	9	100.00	>212.0	>212.0	04/14/2022	15:28
7	N2423-07		1	9	100.00	>212.0	>212.0	04/14/2022	15:45
8	N2423-07DUP		1	10	100.00	>212.0	>212.0	04/14/2022	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 041422 WorkList ID : 158686 Department : Wet-Chemistry Date : 04-14-2022 10:14:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2022	Water	N2389-01	Flash Point	Cool 4 deg C	PSEG03	O11	20564	04/13/2022	1010B
04/20/2022	Water	N2389-03	Flash Point	Cool 4 deg C	PSEG03	O11	33630	04/13/2022	1010B
04/20/2022	Water	N2389-07	Flash Point	Cool 4 deg C	PSEG03	O11	33637-8	04/13/2022	1010B
04/18/2022	Water	N2396-02	Flash Point	Cool 4 deg C	PSEG03	O11	5576	04/13/2022	1010B
04/18/2022	Water	N2396-03	Flash Point	Cool 4 deg C	PSEG03	O11	5581	04/13/2022	1010B
04/21/2022	Water	N2423-07	Flash Point	Cool 4 deg C	PSEG03	O11	D2456-D2459	04/14/2022	1010B

Date/Time 4/14/22 13:30
Raw Sample Received by: AP (WOL)
Raw Sample Relinquished by: AP (WOL)

Date/Time 4/14/22 16:30
Raw Sample Received by: AP (WOL)
Raw Sample Relinquished by: AP (WOL)