



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119691

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	04/14/2022	17:10
2	CAL2	1	Water	NA	NA	20.9	7.00	04/14/2022	17:11
3	CAL3	1	Water	NA	NA	20.9	10.03	04/14/2022	17:12
4	ICV	1	Water	NA	NA	20.9	6.99	04/14/2022	17:15
5	CCV1	1	Water	NA	NA	20.9	2.01	04/14/2022	17:18
6	N2394-02	1	Solid	20.03	20	21.8	8.15	04/14/2022	17:20
7	N2394-02DUP	1	Solid	20.04	20	21.9	8.16	04/14/2022	17:21
8	N2394-04	1	Solid	20.04	20	21.8	9.13	04/14/2022	17:22
9	N2394-06	1	Solid	20.01	20	21.6	8.45	04/14/2022	17:24
10	N2394-08	1	Solid	20.04	20	21.6	8.72	04/14/2022	17:25
11	N2394-10	1	Solid	20.04	20	21.8	8.47	04/14/2022	17:25
12	N2394-12	1	Solid	20.02	20	21.8	8.08	04/14/2022	17:26
13	N2394-14	1	Solid	20.03	20	21.9	8.56	04/14/2022	17:27
14	N2394-16	1	Solid	20.02	20	21.9	9.00	04/14/2022	17:29
15	N2409-01	1	Solid	20.02	20	24.4	6.40	04/14/2022	17:30
16	CCV2	1	Water	NA	NA	20.3	12.03	04/14/2022	17:31
17	N2431-02	1	Solid	20.01	20	24.3	6.89	04/14/2022	17:32
18	CCV3	1	Water	NA	NA	20.2	2.03	04/14/2022	17:36

W3 719691

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2394 WorkList ID : 158701 Department : Wet-Chemistry Date : 04-14-2022 13:07:37

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample		
				Preservative	Customer	Storage Location	Collect Date	Method	
04/13/2022	Solid	N2394-02	Corrosivity	Cool 4 deg C	URSC05	N12	CS-1-COMPOSITE-RIGHT	04/13/2022	9045D
04/13/2022	Solid	N2394-04	Corrosivity	Cool 4 deg C	URSC05	N12	CS-2-COMPOSITE-RIGHT	04/13/2022	9045D
04/13/2022	Solid	N2394-06	Corrosivity	Cool 4 deg C	URSC05	N12	CS-3-SOIL-PILE	04/13/2022	9045D
04/13/2022	Solid	N2394-08	Corrosivity	Cool 4 deg C	URSC05	N12	CS-4-SOIL-PILE	04/13/2022	9045D
04/13/2022	Solid	N2394-10	Corrosivity	Cool 4 deg C	URSC05	N12	CS-5-COMPOSITE-LEFT	04/13/2022	9045D
04/13/2022	Solid	N2394-12	Corrosivity	Cool 4 deg C	URSC05	N12	CS-6-COMPOSITE-LEFT	04/13/2022	9045D
04/13/2022	Solid	N2394-14	Corrosivity	Cool 4 deg C	URSC05	N12	CS-7-SOIL-PILE-BELOW-BRIDGE	04/13/2022	9045D
04/13/2022	Solid	N2394-16	Corrosivity	Cool 4 deg C	URSC05	N12	CS-8	04/13/2022	9045D
04/14/2022	Solid	N2409-01	Corrosivity	Cool 4 deg C	PSEG03	O11	HAR-OILY-DEBRIS	04/14/2022	9045D
04/14/2022	Solid	N2431-02	Corrosivity	Cool 4 deg C	ROUX02	O11	WG-15-1-1	04/14/2022	9045D

Date/Time 04.14.22 16:35
Raw Sample Received by: 2076091
Raw Sample Relinquished by: 2076091

Date/Time 04.14.22 18:00
Raw Sample Received by: L-C (Saw 1)
Raw Sample Relinquished by: 2076091