



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119698

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	83.0	04/15/2022	10:44
2	N2396-01		1	5	18.00	64.4	65.0	04/15/2022	10:59
3	N2396-01DUP		1	5	19.00	66.2	66.8	04/15/2022	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 041522 WorkList ID : 158735 Department : Wet-Chemistry Date : 04-15-2022 10:31:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/18/2022	Solid	N2396-01	Flash Point	Cool 4 deg C	PSEG03	O11	5575	04/13/2022	1010B

Date/Time 4/15/22 10:35
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 4/15/22 11:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]