


LB119699

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.4704e-004

C= -3.4607e-002

Rho= 0.9995913

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.023	-0.023	46	0.000	46					-
0.2	0.200	0.214	0.014	1006	0.0 %	1006					7
2.5	2.500	2.422	-0.078	9943	0.0 %	9943					-3
5.0	5.000	5.059	0.059	20619	0.0 %	20619					1
7.5	7.500	7.682	0.182	31234	0.0 %	31234					2
10.0	10.000	9.847	-0.153	39999	0.0 %	39999					-2

LB119699 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 15 Apr 2022 10:20:59							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.00007470A	PPB		15 Apr 2022 10:25:18	S	Std
0.2 - 1	50.2	1006	-	0.20007470A	PPB		15 Apr 2022 10:27:34	S	Std
2.5 - 1	52.5	9943	-	2.50007470A	PPB		15 Apr 2022 10:29:51	S	Std
5.0 - 1	55	20619	-	5.00007470A	PPB		15 Apr 2022 10:32:08	S	Std
7.5 - 1	57.5	31234	-	7.50007470A	PPB		15 Apr 2022 10:37:01	S	Std
10.0 - 1	57.6	39999	-	10.00007470A	PPB		15 Apr 2022 10:41:51	S	Std
ICV79 - 1	ICV79	16359	4.0068	-7470A	PPB		15 Apr 2022 10:44:35	U	SMPL
ICB79 - 1	ICB79	12	-0.0316	-7470A	PPB		15 Apr 2022 10:49:47	U	SMPL
CCV56 - 1	CCV56	20829	5.1111	-7470A	PPB		15 Apr 2022 10:52:01	U	SMPL
CCB56 - 1	CCB56	2	-0.0341	-7470A	PPB		15 Apr 2022 10:54:16	U	SMPL
CRA - 1	CRA	979	0.2072	-7470A	PPB		15 Apr 2022 10:56:31	U	SMPL
HighStd - 1	HighStd	40550	9.9830	-7470A	PPB		15 Apr 2022 10:58:46	U	SMPL
ChkStd - 1	ChkStd	30557	7.5143	-7470A	PPB		15 Apr 2022 11:01:01	U	SMPL
PB144113BL - 1	PBW	46	-0.0232	-7470A	PPB		15 Apr 2022 11:06:33	U	SMPL
PB144113BS - 1	LCSW	16152	3.9556	-7470A	PPB		15 Apr 2022 11:11:31	U	SMPL
N2364-02 - 1	SB-01-COMP	-63	-0.0502	-7470A	PPB		15 Apr 2022 11:13:47	U	SMPL
N2364-04 - 1	SB-02-COMP	10	-0.0321	-7470A	PPB		15 Apr 2022 11:16:03	U	SMPL
N2364-06 - 1	SB-03-COMP	26	-0.0282	-7470A	PPB		15 Apr 2022 11:18:20	U	SMPL
N2364-08 - 1	SB-09-COMP	22	-0.0292	-7470A	PPB		15 Apr 2022 11:20:37	U	SMPL
N2392-02 - 1	VNJ-212	24	-0.0287	-7470A	PPB		15 Apr 2022 11:22:52	U	SMPL
CCV57 - 1	CCV57	20670	5.0718	-7470A	PPB		15 Apr 2022 11:25:07	U	SMPL
CCB57 - 1	CCB57	24	-0.0287	-7470A	PPB		15 Apr 2022 11:27:23	U	SMPL
N2392-02DUP - 1	VNJ-212DUP	20	-0.0297	-7470A	PPB		15 Apr 2022 11:29:38	U	SMPL
N2392-02MS - 1	VNJ-212MS	17564	4.3045	-7470A	PPB		15 Apr 2022 11:31:54	U	SMPL
N2392-02MSD - 1	VNJ-212MSD	17940	4.3973	-7470A	PPB		15 Apr 2022 11:34:09	U	SMPL
N2396-01 - 1	5575	229	0.0220	-7470A	PPB		15 Apr 2022 11:36:24	U	SMPL
PB144114BL - 1	PBW	24	-0.0287	-7470A	PPB		15 Apr 2022 11:38:41	U	SMPL
PB144114BS - 1	LCSW	16172	3.9606	-7470A	PPB		15 Apr 2022 11:40:57	U	SMPL
N2396-02 - 1	5576	-41	-0.0447	-7470A	PPB		15 Apr 2022 11:43:14	U	SMPL
N2396-03 - 1	5581	8	-0.0326	-7470A	PPB		15 Apr 2022 11:45:30	U	SMPL
N2396-03DUP - 1	5581DUP	30	-0.0272	-7470A	PPB		15 Apr 2022 11:47:47	U	SMPL
N2396-03MS - 1	5581MS	15726	3.8504	-7470A	PPB		15 Apr 2022 11:50:03	U	SMPL
CCV58 - 1	CCV58	20561	5.0448	-7470A	PPB		15 Apr 2022 11:52:18	U	SMPL
CCB58 - 1	CCB58	-30	-0.0420	-7470A	PPB		15 Apr 2022 11:54:34	U	SMPL
N2396-03MSD - 1	5581MSD	15726	3.8504	-7470A	PPB		15 Apr 2022 11:56:49	U	SMPL
PB144154BL - 1	PBW	25	-0.0284	-7470A	PPB		15 Apr 2022 11:59:05	U	SMPL
PB144154BS - 1	LCSW	15995	3.9168	-7470A	PPB		15 Apr 2022 12:01:20	U	SMPL
N2420-02 - 1	RW4-IDWGW-20220413	-69	-0.0517	-7470A	PPB		15 Apr 2022 12:03:35	U	SMPL
N2423-07 - 1	D2456-D2459	30632	7.5328	-7470A	PPB		15 Apr 2022 12:05:51	U	SMPL
N2423-07DUP - 1	D2456-D2459DUP	28817	7.0844	-7470A	PPB		15 Apr 2022 12:10:09	U	SMPL
N2423-07MS - 1	D2456-D2459MS	40422	9.9514	-7470A	PPB		15 Apr 2022 12:12:25	U	SMPL
N2423-07MSD - 1	D2456-D2459MSD	40447	9.9575	-7470A	PPB		15 Apr 2022 12:14:42	U	SMPL
PB144095TB - 1	PB144095TB	-141	-0.0694	-7470A	PPB		15 Apr 2022 12:17:00	U	SMPL
PB144104TB - 1	PB144104TB	14	-0.0311	-7470A	PPB		15 Apr 2022 12:19:17	U	SMPL
CCV59 - 1	CCV59	20331	4.9880	-7470A	PPB		15 Apr 2022 12:21:32	U	SMPL
CCB59 - 1	CCB59	-44	-0.0455	-7470A	PPB		15 Apr 2022 12:23:47	U	SMPL
N2349-20 - 1	COM-1	14	-0.0311	-7470A	PPB		15 Apr 2022 12:27:57	U	SMPL
N2392-02LX5 - 1		-22	-0.0400	-7470A	PPB		15 Apr 2022 12:30:12	U	SMPL
N2392-02A - 1		16079	3.9376	-7470A	PPB		15 Apr 2022 12:32:27	U	SMPL
N2396-03LX5 - 1		-71	-0.0521	-7470A	PPB		15 Apr 2022 12:34:42	U	SMPL
N2396-03A - 1		16390	4.0144	-7470A	PPB		15 Apr 2022 12:36:59	U	SMPL
N2423-07LX5 - 1		5931	1.4306	-7470A	PPB		15 Apr 2022 12:39:15	U	SMPL
N2423-07A - 1		39036	9.6090	-7470A	PPB		15 Apr 2022 12:41:31	U	SMPL
CCV60 - 1	CCV60	19182	4.7042	-7470A	PPB		15 Apr 2022 12:43:48	U	SMPL
CCB60 - 1	CCB60	-25	-0.0408	-7470A	PPB		15 Apr 2022 12:46:06	U	SMPL