



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119702

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	04/15/2022	12:40
2	CAL2	1	Water	NA	NA	20.9	6.99	04/15/2022	12:41
3	CAL3	1	Water	NA	NA	20.9	10.03	04/15/2022	12:43
4	ICV	1	Water	NA	NA	20.8	7.01	04/15/2022	12:45
5	CCV1	1	Water	NA	NA	20.8	2.01	04/15/2022	12:46
6	N2441-02	1	Water	NA	NA	20.1	3.24	04/15/2022	12:50
7	N2441-02DUP	1	Water	NA	NA	20.1	3.25	04/15/2022	12:51
8	CCV2	1	Water	NA	NA	20.8	12.03	04/15/2022	12:53