

LB119710


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57079



A= 0.0000e+000

B= 1.7492e-004

C= -3.7926e-002

Rho= 0.9999347

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.027	-0.027	62	0.000	62					
0.2	0.200	0.202	0.002	1370	0.0 %	1370					
2.5	2.500	2.524	0.024	14649	0.0 %	14649					
5.0	5.000	4.978	-0.022	28675	0.0 %	28675					
7.5	7.500	7.577	0.077	43535	0.0 %	43535					
10.0	10.000	9.946	-0.054	57079	0.0 %	57079					

LB119710 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Apr 2022 14:49:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	62	-	0.0000	7471B	PPB	15 Apr 2022 14:52:34	S	Std
0.2 - 1	50.2	1370	-	0.2000	7471B	PPB	15 Apr 2022 14:54:51	S	Std
2.5 - 1	52.5	14649	-	2.5000	7471B	PPB	15 Apr 2022 14:57:08	S	Std
5.0 - 1	55	28675	-	5.0000	7471B	PPB	15 Apr 2022 14:59:24	S	Std
7.5 - 1	57.5	43535	-	7.5000	7471B	PPB	15 Apr 2022 15:01:41	S	Std
10.0 - 1	511	57079	-	10.0000	7471B	PPB	15 Apr 2022 15:03:57	S	Std
ICV80 - 1	ICV80	21236	3.6766		7471B	PPB	15 Apr 2022 15:06:41	U	SMPL
ICB80 - 1	ICB80	-66	-0.0495		7471B	PPB	15 Apr 2022 15:08:56	U	SMPL
CCV81 - 1	CCV81	28819	5.0030		7471B	PPB	15 Apr 2022 15:11:11	U	SMPL
CCB81 - 1	CCB81	-80	-0.0519		7471B	PPB	15 Apr 2022 15:13:26	U	SMPL
CRA - 1	CRA	1337	0.1959		7471B	PPB	15 Apr 2022 15:15:42	U	SMPL
HighStd - 1	HighStd	52404	9.1284		7471B	PPB	15 Apr 2022 15:30:13	U	SMPL
ChkStd - 1	ChkStd	37534	6.5274		7471B	PPB	15 Apr 2022 15:32:30	U	SMPL
PB144162BL - 1	PBS	-271	-0.0853		7471B	PPB	15 Apr 2022 15:34:50	U	SMPL
PB144162BS - 1	LCSS	21469	3.7173		7471B	PPB	15 Apr 2022 15:42:15	U	SMPL
N2335-02 - 1	A350-CU	14763	2.5444		7471B	PPB	15 Apr 2022 15:44:31	U	SMPL
N2336-02 - 1	C150-CU	4155	0.6889		7471B	PPB	15 Apr 2022 15:46:50	U	SMPL
N2337-02 - 1	B250-CU	2035	0.3180		7471B	PPB	15 Apr 2022 15:49:08	U	SMPL
N2390-01 - 1	CL-1-041422	3519	0.5776		7471B	PPB	15 Apr 2022 15:56:12	U	SMPL
N2390-01DUP - 1	CL-1-041422DUP	3555	0.5839		7471B	PPB	15 Apr 2022 15:58:27	U	SMPL
CCV82 - 1	CCV82	28093	4.8760		7471B	PPB	15 Apr 2022 16:05:58	U	SMPL
CCB82 - 1	CCB82	-152	-0.0645		7471B	PPB	15 Apr 2022 16:08:12	U	SMPL
N2390-01MS - 1	CL-1-041422MS	28355	4.9218		7471B	PPB	15 Apr 2022 16:10:30	U	SMPL
N2390-01MSD - 1	CL-1-041422MSD	27486	4.7698		7471B	PPB	15 Apr 2022 16:12:45	U	SMPL
N2392-01 - 1	VNJ-212	386	0.0296		7471B	PPB	15 Apr 2022 16:15:02	U	SMPL
N2409-01 - 1	HAR-OILY-DEBRIS	104	-0.0197		7471B	PPB	15 Apr 2022 16:17:20	U	SMPL
N2423-05 - 1	D2460-D2462	521	0.0532		7471B	PPB	15 Apr 2022 16:19:35	U	SMPL
N2424-01 - 1	OR-02-041422	4116	0.6820		7471B	PPB	15 Apr 2022 16:21:52	U	SMPL
N2390-01LX5 - 1		717	0.0875		7471B	PPB	15 Apr 2022 16:24:08	U	SMPL
N2390-01A - 1		25344	4.3951		7471B	PPB	15 Apr 2022 16:26:25	U	SMPL
CCV83 - 1	CCV83	27832	4.8303		7471B	PPB	15 Apr 2022 16:28:42	U	SMPL
CCB83 - 1	CCB83	-72	-0.0505		7471B	PPB	15 Apr 2022 16:30:59	U	SMPL