



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119711

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/15/2022	15:40
2	CAL2	1	Water	NA	NA	20.8	7.00	04/15/2022	15:41
3	CAL3	1	Water	NA	NA	20.8	10.03	04/15/2022	15:42
4	ICV	1	Water	NA	NA	20.8	6.99	04/15/2022	15:45
5	CCV1	1	Water	NA	NA	20.9	2.02	04/15/2022	15:47
6	N2449-01	1	Solid	20.02	20	22.0	4.43	04/15/2022	15:50
7	N2449-02	1	Solid	20.03	20	22.1	6.31	04/15/2022	15:51
8	N2449-02DUP	1	Solid	20.04	20	22.2	6.32	04/15/2022	15:52
9	CCV2	1	Water	NA	NA	20.4	12.03	04/15/2022	16:00

23119711

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S N2449 WorkList ID : 158746 Department : Wet-Chemistry Date : 04-15-2022 15:30:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/15/2022	Solid	N2449-01	pH	Cool 4 deg C	ENV181	O11	22L091-A	04/15/2022	9045D
04/15/2022	Solid	N2449-02	pH	Cool 4 deg C	ENV181	O11	22L091-B	04/15/2022	9045D

Date/Time 04/15/22 15:35
Raw Sample Received by: DP/WR
Raw Sample Relinquished by: CP/SM

Date/Time 04/15/22 17:00
Raw Sample Received by: CP/SM
Raw Sample Relinquished by: DP/WR