



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119723

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/15/2022	17:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/15/2022	17:16
3	CAL3	1	Water	NA	NA	20.3	10.03	04/15/2022	17:18
4	ICV	1	Water	NA	NA	20.4	6.99	04/15/2022	17:20
5	CCV1	1	Water	NA	NA	20.4	2.01	04/15/2022	17:22
6	N2439-02	1	Solid	20.03	20	20.1	5.14	04/15/2022	17:24
7	N2439-04	1	Solid	20.04	20	22.5	6.40	04/15/2022	17:25
8	N2452-02	1	Solid	20.02	20	21.7	5.63	04/15/2022	17:30
9	N2452-06	1	Solid	20.03	20	21.6	6.84	04/15/2022	17:33
10	N2452-10	1	Solid	20.04	20	21.8	6.86	04/15/2022	17:35
11	N2452-14	1	Solid	20.02	20	21.7	9.43	04/15/2022	17:36
12	N2452-14DUP	1	Solid	20.03	20	21.8	9.44	04/15/2022	17:37
13	CCV2	1	Water	NA	NA	20.7	12.03	04/15/2022	17:40

87119423

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2434 WorkList ID : 158749 Department : Wet-Chemistry Date : 04-15-2022 15:51:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/15/2022	Solid	N2452-02	Corrosivity	Cool 4 deg C	ROUX02	O11	WC-15-1-2	04/15/2022	9045D
04/15/2022	Solid	N2452-06	Corrosivity	Cool 4 deg C	ROUX02	O11	WC-15-1-3	04/15/2022	9045D
04/15/2022	Solid	N2452-10	Corrosivity	Cool 4 deg C	ROUX02	O11	WC-15-1-4	04/15/2022	9045D
04/15/2022	Solid	N2452-14	Corrosivity	Cool 4 deg C	ROUX02	O11	WC-15-1-5	04/15/2022	9045D
04/15/2022	Solid	N2439-02	Corrosivity	Cool 4 deg C	PSEG03	O11	WARD-OIL	04/15/2022	9045D
04/15/2022	Solid	N2439-04	Corrosivity	Cool 4 deg C	PSEG03	O11	WARD-OD	04/15/2022	9045D

Date/Time 04.15.22 16:20
 Raw Sample Received by: 70(CDC)
 Raw Sample Relinquished by: 70(CDC)

Date/Time 04.15.22 18:00
 Raw Sample Received by: 12(CDC)
 Raw Sample Relinquished by: 20(CDC)