


L13119724

7470A

INSTRUMENT ID: 011

μ Abs.:

40619



Linear

A= 0.0000e+000

B= 2.4515e-004

C= -2.5417e-002

Rho= 0.9999154

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	-0.034	-0.034	-37	0.000	-37					
0.2	0.200	0.208	0.008	951	0.0 %	951					4
2.5	2.500	2.475	-0.025	10198	0.0 %	10198					+1
5.0	5.000	5.070	0.070	20783	0.0 %	20783					1
7.5	7.500	7.550	0.050	30902	0.0 %	30902					1
10.0	10.000	9.932	-0.068	40619	0.0 %	40619					-1

LB119724 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Apr 2022 10:57:13

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-37	-	0.0007470A	PPB	18 Apr 2022 11:07:09	S	Std
0.2 - 1	50.2	951	-	0.20007470A	PPB	18 Apr 2022 11:09:25	S	Std
2.5 - 1	52.5	10198	-	2.50007470A	PPB	18 Apr 2022 11:11:42	S	Std
5.0 - 1	55	20783	-	5.00007470A	PPB	18 Apr 2022 11:13:59	S	Std
7.5 - 1	57.5	30902	-	7.50007470A	PPB	18 Apr 2022 11:16:15	S	Std
10.0 - 1	59.5	40619	-	10.00007470A	PPB	18 Apr 2022 11:23:35	S	Std
ICV81 - 1	ICV81	15162	3.6916	-7470A	PPB	18 Apr 2022 11:26:24	U	SMPL
ICB81 - 1	ICB81	-43	-0.0360	-7470A	PPB	18 Apr 2022 11:28:39	U	SMPL
CCV84 - 1	CCV84	20418	4.9801	-7470A	PPB	18 Apr 2022 11:30:54	U	SMPL
CCB84 - 1	CCB84	-51	-0.0379	-7470A	PPB	18 Apr 2022 11:33:08	U	SMPL
CRA - 1	CRA	867	0.1871	-7470A	PPB	18 Apr 2022 11:35:23	U	SMPL
HighStd - 1	HighStd	39747	9.7188	-7470A	PPB	18 Apr 2022 11:37:38	U	SMPL
ChkStd - 1	ChkStd	28394	6.9354	-7470A	PPB	18 Apr 2022 11:39:53	U	SMPL
PB144157BL - 1	PBW	6	-0.0239	-7470A	PPB	18 Apr 2022 11:44:43	U	SMPL
PB144157BS - 1	LCSW	16908	4.1196	-7470A	PPB	18 Apr 2022 11:46:58	U	SMPL
N2255-02 - 1	31495	445	0.0837	-7470A	PPB	18 Apr 2022 11:49:14	U	SMPL
N2394-02 - 1	CS-1(DUMPSTER-RIGHT)	-46	-0.0367	-7470A	PPB	18 Apr 2022 11:51:31	U	SMPL
N2394-02DUP - 1	CS-1(DUMPSTER-RIGHT)DUP	20	-0.0205	-7470A	PPB	18 Apr 2022 11:53:48	U	SMPL
N2394-02MS - 1	CS-1(DUMPSTER-RIGHT)MS	17676	4.3079	-7470A	PPB	18 Apr 2022 11:56:04	U	SMPL
N2394-02MSD - 1	CS-1(DUMPSTER-RIGHT)MSD	16559	4.0340	-7470A	PPB	18 Apr 2022 11:58:20	U	SMPL
CCV85 - 1	CCV85	20637	5.0338	-7470A	PPB	18 Apr 2022 12:00:35	U	SMPL
CCB85 - 1	CCB85	-89	-0.0472	-7470A	PPB	18 Apr 2022 12:02:50	U	SMPL
N2394-04 - 1	CS-2(DUMPSTER-RIGHT)	1	-0.0252	-7470A	PPB	18 Apr 2022 12:05:05	U	SMPL
N2394-06 - 1	CS-3-SOIL-PILE	8	-0.0235	-7470A	PPB	18 Apr 2022 12:07:20	U	SMPL
N2394-08 - 1	CS-4-SOIL-PILE	13	-0.0222	-7470A	PPB	18 Apr 2022 12:09:36	U	SMPL
N2394-10 - 1	CS-5(DUMPSTER-LEFT)	-25	-0.0315	-7470A	PPB	18 Apr 2022 12:11:51	U	SMPL
N2394-12 - 1	CS-6(DUMPSTER-LEFT)	3	-0.0247	-7470A	PPB	18 Apr 2022 12:14:07	U	SMPL
N2394-14 - 1	CS-7-SOIL-PILE-BELOW-BRIDGE	107	0.0008	-7470A	PPB	18 Apr 2022 12:16:25	U	SMPL
N2394-16 - 1	CS-8	30	-0.0181	-7470A	PPB	18 Apr 2022 12:18:41	U	SMPL
N2431-02 - 1	WC-15-1-1	51	-0.0129	-7470A	PPB	18 Apr 2022 12:20:57	U	SMPL
PB144182BL - 1	PBW	21	-0.0203	-7470A	PPB	18 Apr 2022 12:23:14	U	SMPL
PB144182BS - 1	LCSW	14508	3.5312	-7470A	PPB	18 Apr 2022 12:25:29	U	SMPL
CCV86 - 1	CCV86	20346	4.9624	-7470A	PPB	18 Apr 2022 12:27:44	U	SMPL
CCB86 - 1	CCB86	-83	-0.0458	-7470A	PPB	18 Apr 2022 12:30:00	U	SMPL
N2432-05 - 1	FB220414	-20	-0.0303	-7470A	PPB	18 Apr 2022 12:32:15	U	SMPL
N2441-02 - 1	TT-041-IDWGW-20220413	-7	-0.0271	-7470A	PPB	18 Apr 2022 12:34:30	U	SMPL
N2448-01 - 1	E14ST-B6-041522-GW	23	-0.0198	-7470A	PPB	18 Apr 2022 12:36:45	U	SMPL
N2448-01DUP - 1	E14ST-B6-041522-GWDUP	18	-0.0210	-7470A	PPB	18 Apr 2022 12:39:01	U	SMPL
N2448-01MS - 1	E14ST-B6-041522-GWMS	15352	3.7381	-7470A	PPB	18 Apr 2022 12:41:17	U	SMPL
N2448-01MSD - 1	E14ST-B6-041522-GWMSD	16580	4.0392	-7470A	PPB	18 Apr 2022 12:43:33	U	SMPL
N2448-02 - 1	E14ST-B6-041522-GW-DUP	-16	-0.0293	-7470A	PPB	18 Apr 2022 12:45:49	U	SMPL
CCV87 - 1	CCV87	20277	4.9455	-7470A	PPB	18 Apr 2022 12:48:05	U	SMPL
CCB87 - 1	CCB87	-61	-0.0404	-7470A	PPB	18 Apr 2022 12:50:22	U	SMPL
N2394-08LX5 - 1	PB144157TB	8	-0.0235	-7470A	PPB	18 Apr 2022 12:52:38	U	SMPL
N2394-08A - 1	N2394-02LX5	-3	-0.0262	-7470A	PPB	18 Apr 2022 12:54:53	U	SMPL
N2448-01LX5 - 1	N2394-02A	17189	4.1885	-7470A	PPB	18 Apr 2022 12:57:08	U	SMPL
N2448-01A - 1		-81	-0.0453	-7470A	PPB	18 Apr 2022 12:59:24	U	SMPL
CCV88 - 1	CCV88	16461	4.0100	-7470A	PPB	18 Apr 2022 13:01:39	U	SMPL
CCB88 - 1	CCB88	20440	4.9855	-7470A	PPB	18 Apr 2022 13:06:09	U	SMPL
		-73	-0.0433	-7470A	PPB	18 Apr 2022 13:08:24	U	SMPL