

L3119729


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

56767



A= 0.0000e+000

B= 1.7655e-004

C= 1.7776e-002

Rho= 0.9999464

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.025	0.025	40	0.000	40					
0.2	0.200	0.234	0.034	1223	0.0 %	1223					
2.5	2.500	2.469	-0.031	13881	0.0 %	13881					
5.0	5.000	4.934	-0.066	27844	0.0 %	27844					
7.5	7.500	7.499	-0.001	42373	0.0 %	42373					
10.0	10.000	10.040	0.040	56767	0.0 %	56767					

100
100
100
100
100

LB119729 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 18 Apr 2022 14:47:51					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.00007471B	PPB		18 Apr 2022 14:50:30	S	Std
0.2 - 1	52	1223	-	0.20007471B	PPB		18 Apr 2022 14:52:47	S	Std
2.5 - 1	52.5	13881	-	2.50007471B	PPB		18 Apr 2022 14:55:04	S	Std
5.0 - 1	55	27844	-	5.00007471B	PPB		18 Apr 2022 14:57:20	S	Std
7.5 - 1	57.5	42373	-	7.50007471B	PPB		18 Apr 2022 14:59:36	S	Std
10.0 - 1	59	56767	-	10.00007471B	PPB		18 Apr 2022 15:01:53	S	Std
ICV82 - 1	ICV82	21180	3.7572	-7471B	PPB		18 Apr 2022 15:04:49	U	SMPL
ICB82 - 1	ICB82	-90	0.0019	-7471B	PPB		18 Apr 2022 15:07:03	U	SMPL
CCV89 - 1	CCV89	28480	5.0460	-7471B	PPB		18 Apr 2022 15:09:20	U	SMPL
CCB89 - 1	CCB89	-89	0.0021	-7471B	PPB		18 Apr 2022 15:11:35	U	SMPL
CRA - 1	CRA	1216	0.2325	-7471B	PPB		18 Apr 2022 15:13:51	U	SMPL
HighStd - 1	HighStd	57549	10.1783	-7471B	PPB		18 Apr 2022 15:16:06	U	SMPL
ChkStd - 1	ChkStd	39206	6.9398	-7471B	PPB		18 Apr 2022 15:18:21	U	SMPL
PB144197BL - 1	PBS	-126	-0.0045	-7471B	PPB		18 Apr 2022 15:20:38	U	SMPL
PB144197BS - 1	LCSS	21078	3.7392	-7471B	PPB		18 Apr 2022 15:22:55	U	SMPL
N2391-01 - 1	TR-06-041422	1699	0.3177	-7471B	PPB		18 Apr 2022 15:30:01	U	SMPL
N2391-01DUP - 1	TR-06-041422DUP	1678	0.3140	-7471B	PPB		18 Apr 2022 15:32:17	U	SMPL
N2391-01MS - 1	TR-06-041422MS	23152	4.1054	-7471B	PPB		18 Apr 2022 15:34:34	U	SMPL
N2391-01MSD - 1	TR-06-041422MSD	25015	4.4343	-7471B	PPB		18 Apr 2022 15:36:51	U	SMPL
N2429-01 - 1	BU-03-041422	2567	0.4710	-7471B	PPB		18 Apr 2022 15:39:06	U	SMPL
CCV90 - 1	CCV90	28092	4.9775	-7471B	PPB		18 Apr 2022 15:41:21	U	SMPL
CCB90 - 1	CCB90	-49	0.0091	-7471B	PPB		18 Apr 2022 15:43:37	U	SMPL
N2430-02 - 1	E13-EB1-041322-0-5	15197	2.7009	-7471B	PPB		18 Apr 2022 15:45:53	U	SMPL
N2430-03 - 1	E13-EB1-041322-5-10	662	0.1347	-7471B	PPB		18 Apr 2022 15:48:08	U	SMPL
N2432-01 - 1	SB-6075	549	0.1147	-7471B	PPB		18 Apr 2022 15:50:23	U	SMPL
N2432-02 - 1	SB-5115	1196	0.2289	-7471B	PPB		18 Apr 2022 15:52:39	U	SMPL
N2432-03 - 1	SB-4115	347	0.0790	-7471B	PPB		18 Apr 2022 15:54:54	U	SMPL
N2432-04 - 1	SB-2105	607	0.1249	-7471B	PPB		18 Apr 2022 15:57:11	U	SMPL
N2432-08 - 1	DUP220414	2828	0.5171	-7471B	PPB		18 Apr 2022 15:59:27	U	SMPL
N2434-01 - 1	RCA-EDISON-PILE-1	1437	0.2715	-7471B	PPB		18 Apr 2022 16:01:47	U	SMPL
N2435-01 - 1	BBC	5647	1.0148	-7471B	PPB		18 Apr 2022 16:04:03	U	SMPL
N2442-01 - 1	124035	5692	1.0227	-7471B	PPB		18 Apr 2022 16:06:19	U	SMPL
CCV91 - 1	CCV91	28341	5.0215	-7471B	PPB		18 Apr 2022 16:08:34	U	SMPL
CCB91 - 1	CCB91	-31	0.0123	-7471B	PPB		18 Apr 2022 16:10:50	U	SMPL
N2442-03 - 1	124030	5777	1.0377	-7471B	PPB		18 Apr 2022 16:13:06	U	SMPL
N2442-05 - 1	124042	2314	0.4263	-7471B	PPB		18 Apr 2022 16:15:21	U	SMPL
N2445-01 - 1	282	1201	0.2298	-7471B	PPB		18 Apr 2022 16:17:36	U	SMPL
N2446-01 - 1	339	2093	0.3873	-7471B	PPB		18 Apr 2022 16:19:52	U	SMPL
N2449-01 - 1	22L091-A	2577	0.4728	-7471B	PPB		18 Apr 2022 16:22:08	U	SMPL
N2449-02 - 1	22L091-B	3037	0.5540	-7471B	PPB		18 Apr 2022 16:24:25	U	SMPL
N2391-01LX5 - 1		295	0.0699	-7471B	PPB		18 Apr 2022 16:26:41	U	SMPL
N2391-01A - 1		24466	4.3374	-7471B	PPB		18 Apr 2022 16:28:57	U	SMPL
CCV92 - 1	CCV92	28439	5.0388	-7471B	PPB		18 Apr 2022 16:31:14	U	SMPL
CCB92 - 1	CCB92	-41	0.0105	-7471B	PPB		18 Apr 2022 16:33:30	U	SMPL