



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119734

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/18/2022	17:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/18/2022	17:16
3	CAL3	1	Water	NA	NA	20.3	10.03	04/18/2022	17:19
4	ICV	1	Water	NA	NA	20.4	7.00	04/18/2022	17:20
5	CCV1	1	Water	NA	NA	20.4	2.01	04/18/2022	17:22
6	N2459-01	1	Solid	20.03	20	20.5	7.86	04/18/2022	17:25
7	N2459-02	1	Solid	20.04	20	20.3	7.79	04/18/2022	17:26
8	N2468-04	1	Solid	20.02	20	20.8	7.88	04/18/2022	17:30
9	N2468-08	1	Solid	20.03	20	20.7	7.93	04/18/2022	17:32
10	N2468-12	1	Solid	20.04	20	20.7	7.93	04/18/2022	17:35
11	N2468-12DUP	1	Solid	20.03	20	20.6	7.80	04/18/2022	17:36
12	CCV2	1	Water	NA	NA	20.3	12.03	04/18/2022	17:40

10719434

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2468 WorkList ID : 158786 Department : Wet-Chemistry Date : 04-18-2022 16:52:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
04/18/2022	Solid	N2459-01	Corrosivity	Cool 4 deg C	PSEG03	P11	341	04/18/2022	9045D
04/18/2022	Solid	N2459-02	Corrosivity	Cool 4 deg C	PSEG03	P11	201	04/18/2022	9045D
04/18/2022	Solid	N2468-04	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-15-1-6	04/18/2022	9045D
04/18/2022	Solid	N2468-08	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-15-1-7	04/18/2022	9045D
04/18/2022	Solid	N2468-12	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-15-1-8	04/18/2022	9045D

Date/Time 04-18-22 16:35
 Raw Sample Received by: JD (WCI)
 Raw Sample Relinquished by: JTCM

Date/Time 04-18-22 18:00
 Raw Sample Received by: JTCM
 Raw Sample Relinquished by: JD (WCI)