



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119736

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	04/19/2022	08:14
2	CAL2	1	Water	NA	NA	20.1	7.00	04/19/2022	08:15
3	CAL3	1	Water	NA	NA	20.1	10.02	04/19/2022	08:16
4	ICV	1	Water	NA	NA	20.2	6.99	04/19/2022	08:19
5	CCV1	1	Water	NA	NA	20.3	2.01	04/19/2022	08:20
6	N2459-03	1	Water	NA	NA	20.1	6.82	04/19/2022	08:22
7	N2459-04	1	Water	NA	NA	20.6	6.71	04/19/2022	08:23
8	N2459-05	1	Water	NA	NA	20.4	6.95	04/19/2022	08:25
9	N2459-06	1	Water	NA	NA	20.3	6.82	04/19/2022	08:26
10	N2459-07	1	Water	NA	NA	20.8	6.88	04/19/2022	08:30
11	N2459-08	1	Water	NA	NA	20.3	7.25	04/19/2022	08:32
12	N2459-09	1	Water	NA	NA	20.1	7.19	04/19/2022	08:33
13	N2459-10	1	Water	NA	NA	20.7	7.69	04/19/2022	08:35
14	N2459-11	1	Water	NA	NA	20.4	7.59	04/19/2022	08:38
15	N2459-12	1	Water	NA	NA	20.5	7.51	04/19/2022	08:40
16	CCV2	1	Water	NA	NA	20.5	12.03	04/19/2022	08:42
17	N2459-13	1	Water	NA	NA	20.8	7.66	04/19/2022	08:45
18	N2459-14	1	Water	NA	NA	20.3	7.55	04/19/2022	08:47
19	N2459-18	1	Water	NA	NA	20.1	6.72	04/19/2022	08:50
20	N2459-20	1	Water	NA	NA	20.8	6.95	04/19/2022	08:52

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N2462-05	1	Water	NA	NA	20.2	6.34	04/19/2022	08:55
22	N2462-05DUP	1	Water	NA	NA	20.3	6.35	04/19/2022	08:56
23	CCV3	1	Water	NA	NA	20.4	2.01	04/19/2022	09:00

114736

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2459 WorkList ID : 158791 Department : Wet-Chemistry Date : 04-19-2022 08:06:59

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
04/18/2022	Water	N2459-03	pH	Cool 4 deg C	PSEG03	P11	31573	04/18/2022	9040C
04/18/2022	Water	N2459-04	pH	Cool 4 deg C	PSEG03	P11	31574	04/18/2022	9040C
04/18/2022	Water	N2459-05	pH	Cool 4 deg C	PSEG03	P11	31575	04/18/2022	9040C
04/18/2022	Water	N2459-06	pH	Cool 4 deg C	PSEG03	P11	31576	04/18/2022	9040C
04/18/2022	Water	N2459-07	pH	Cool 4 deg C	PSEG03	P11	31577	04/18/2022	9040C
04/18/2022	Water	N2459-08	pH	Cool 4 deg C	PSEG03	P11	31578	04/18/2022	9040C
04/18/2022	Water	N2459-09	pH	Cool 4 deg C	PSEG03	P11	31579	04/18/2022	9040C
04/18/2022	Water	N2459-10	pH	Cool 4 deg C	PSEG03	P11	31580	04/18/2022	9040C
04/18/2022	Water	N2459-11	pH	Cool 4 deg C	PSEG03	P11	31581	04/18/2022	9040C
04/18/2022	Water	N2459-12	pH	Cool 4 deg C	PSEG03	P11	31582	04/18/2022	9040C
04/18/2022	Water	N2459-13	pH	Cool 4 deg C	PSEG03	P11	31583	04/18/2022	9040C
04/18/2022	Water	N2459-14	pH	Cool 4 deg C	PSEG03	P11	31584	04/18/2022	9040C
04/18/2022	Water	N2459-18	pH	Cool 4 deg C	PSEG03	P11	4922A	04/18/2022	9040C
04/18/2022	Water	N2459-20	pH	Cool 4 deg C	PSEG03	P11	4922B	04/18/2022	9040C
04/18/2022	Water	N2462-05	pH	Cool 4 deg C	PSEG03	P11	295-OW	04/18/2022	9040C

04-19-22 08:10

Date/Time
Raw Sample Received by: CP
Raw Sample Relinquished by: CP

04-18-22 09:40

Date/Time
Raw Sample Received by: CP
Raw Sample Relinquished by: CP