


LB119740

7471B

INSTRUMENT ID: CV1

 μ Abs.:

57123



Linear

A= 0.0000e+000

B= 1.7296e-004

C= -1.4081e-002

Rho= 0.9996536

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	12	0.000	12					-
0.2	0.200	0.193	-0.007	1198	0.0 %	1198					-4
2.5	2.500	2.485	-0.015	14451	0.0 %	14451					-1
5.0	5.000	4.974	-0.026	28839	0.0 %	28839					-1
7.5	7.500	7.694	0.194	44564	0.0 %	44564					3
10.0	10.000	9.866	-0.134	57123	0.0 %	57123					-1

LB119740 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Apr 2022 10:03:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7471B	PPB	19 Apr 2022 10:22:57	S	Std
0.2 - 1	50.2	1198	-	0.2000	7471B	PPB	19 Apr 2022 10:25:14	S	Std
2.5 - 1	52.5	14451	-	2.5000	7471B	PPB	19 Apr 2022 10:27:31	S	Std
5.0 - 1	55	28839	-	5.0000	7471B	PPB	19 Apr 2022 10:29:48	S	Std
7.5 - 1	57.5	44564	-	7.5000	7471B	PPB	19 Apr 2022 10:32:05	S	Std
10.0 - 1	57.0	57123	-	10.0000	7471B	PPB	19 Apr 2022 10:34:20	S	Std
ICV83 - 1	ICV83	22069	3.8030	-	7471B	PPB	19 Apr 2022 10:37:02	U	SMPL
ICB83 - 1	ICB83	-71	-0.0264	-	7471B	PPB	19 Apr 2022 10:39:16	U	SMPL
CCV93 - 1	CCV93	29365	5.0649	-	7471B	PPB	19 Apr 2022 10:41:31	U	SMPL
CCB93 - 1	CCB93	-78	-0.0276	-	7471B	PPB	19 Apr 2022 10:43:46	U	SMPL
CRA - 1	CRA	1275	0.2064	-	7471B	PPB	19 Apr 2022 10:46:02	U	SMPL
HighStd - 1	HighStd	57757	9.9756	-	7471B	PPB	19 Apr 2022 10:48:17	U	SMPL
ChkStd - 1	ChkStd	41418	7.1496	-	7471B	PPB	19 Apr 2022 10:50:33	U	SMPL
PB144224BL - 1	PB144224BL	-137	-0.0378	-	7471B	PPB	19 Apr 2022 10:52:50	U	SMPL
PB144224BS - 1	PB144224BS	24531	4.2288	-	7471B	PPB	19 Apr 2022 10:55:07	U	SMPL
N2451-01 - 1	SB-8085	623	0.0937	-	7471B	PPB	19 Apr 2022 10:57:23	U	SMPL
N2451-02 - 1	SB-7085	564	0.0835	-	7471B	PPB	19 Apr 2022 10:59:39	U	SMPL
N2451-03 - 1	SB-9145	530	0.0776	-	7471B	PPB	19 Apr 2022 11:01:56	U	SMPL
N2451-04 - 1	SB-3145	380	0.0516	-	7471B	PPB	19 Apr 2022 11:04:12	U	SMPL
N2451-05 - 1	SB-10145	123	0.0072	-	7471B	PPB	19 Apr 2022 11:06:28	U	SMPL
CCV94 - 1	CCV94	29374	5.0664	-	7471B	PPB	19 Apr 2022 11:08:43	U	SMPL
CCB94 - 1	CCB94	-62	-0.0248	-	7471B	PPB	19 Apr 2022 11:10:57	U	SMPL
N2451-06 - 1	SB-1120	400	0.0551	-	7471B	PPB	19 Apr 2022 11:13:12	U	SMPL
N2462-01 - 1	296-05	579	0.0861	-	7471B	PPB	19 Apr 2022 11:15:27	U	SMPL
N2462-03 - 1	COMP-05	3056	0.5145	-	7471B	PPB	19 Apr 2022 11:17:42	U	SMPL
N2463-01 - 1	HD-01-041822	2030	0.3370	-	7471B	PPB	19 Apr 2022 11:19:58	U	SMPL
N2467-01 - 1	KR-1	1051	0.1677	-	7471B	PPB	19 Apr 2022 11:27:03	U	SMPL
N2467-01DUP - 1	KR-1DUP	1049	0.1674	-	7471B	PPB	19 Apr 2022 11:33:28	U	SMPL
N2467-01MS - 1	KR-1MS	24830	4.2805	-	7471B	PPB	19 Apr 2022 11:35:44	U	SMPL
N2467-01MSD - 1	KR-1MSD	23669	4.0797	-	7471B	PPB	19 Apr 2022 11:38:01	U	SMPL
N2467-01LX5 - 1		142	0.0105	-	7471B	PPB	19 Apr 2022 11:40:17	U	SMPL
N2467-01A - 1		24065	4.1482	-	7471B	PPB	19 Apr 2022 11:42:33	U	SMPL
CCV95 - 1	CCV95	29170	5.0312	-	7471B	PPB	19 Apr 2022 11:44:49	U	SMPL
CCB95 - 1	CCB95	-81	-0.0281	-	7471B	PPB	19 Apr 2022 11:47:04	U	SMPL