

=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

4/19/2022 11:02

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.459	0.0	0.083	
ICB1	-0.096	0.0	0.002	
CCV1	242.601	0.0	0.210	
CCB1	-0.178	0.0	0.002	
PB144198BL	-0.200	0.0	0.002	
N2399-02	0.574	0.0	0.002	
N2399-02DUP	0.476	0.0	0.002	
N2400-02	0.298	0.0	0.002	
CCV2	244.759	0.0	0.212	
CCB2	-0.378	0.0	0.001	

N	10
Mean	58.232
SD	102.0765
CV%	175.29

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 19 10:11:55 2022

Tue Apr 19 11:01:34 2022

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.4997	µg/l	4/19/2022 10:11:55	
5.OPPBCN	A	Total CN	P	4.7476	µg/l	4/19/2022 10:11:56	
10PPBCN	A	Total CN	P	9.4876	µg/l	4/19/2022 10:11:57	
50PPBCN	A	Total CN	P	49.6497	µg/l	4/19/2022 10:11:58	
100PPBCN	A	Total CN	P	100.5198	µg/l	4/19/2022 10:11:59	
250PPBCN	A	Total CN	P	252.3023	µg/l	4/19/2022 10:12:00	
500PPBCN	A	Total CN	P	498.7927	µg/l	4/19/2022 10:12:01	
ICV1	S	Total CN	P	94.4587	µg/l	4/19/2022 10:37:29	
ICB1	S	Total CN	P	-0.0958	µg/l	4/19/2022 10:37:32	
CCV1	S	Total CN	P	242.6009	µg/l	4/19/2022 10:37:33	
CCB1	S	Total CN	P	-0.1775	µg/l	4/19/2022 10:37:36	
PB144198BL	S	Total CN	P	-0.2001	µg/l	4/19/2022 10:37:38	
N2399-02	S	Total CN	P	0.5736	µg/l	4/19/2022 10:37:39	
N2399-02DUP	S	Total CN	P	0.4758	µg/l	4/19/2022 10:41:48	
N2400-02	S	Total CN	P	0.2983	µg/l	4/19/2022 11:01:31	
CCV2	S	Total CN	P	244.7595	µg/l	4/19/2022 11:01:33	
CCB2	S	Total CN	P	-0.3776	µg/l	4/19/2022 11:01:34	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

4/19/2022 10:13


Test Total CN

Accepted 4/19/2022 10:13

Factor 1162
Bias 0.002

Coeff. of det. 0.999962

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.4997	0.0000	-
2	5.0PPBCN	0.006	4.7476	5.0000	-5.0
3	10PPBCN	0.010	9.4876	10.0000	-5.1
4	50PPBCN	0.044	49.6497	50.0000	-0.7
5	100PPBCN	0.088	100.5198	100.0000	0.5
6	250PPBCN	0.219	252.3023	250.0000	0.9
7	500PPBCN	0.431	498.7927	500.0000	-0.3

12
4/19/22