



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB119744

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	04/18/2022	17:05
2	N2439-02		1	10	100.00	>212.0	>212.0	04/18/2022	17:25
3	N2439-02DUP		1	10	100.00	>212.0	>212.0	04/18/2022	17:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 11 9744

WorkList Name : FP41822 WorkList ID : 158785 Department : Wet-Chemistry Date : 04-18-2022 16:35:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2022	Solid	N2439-02	Flash Point	Cool 4 deg C	PSEG03	O11	WARD-OIL	04/15/2022	1010B

Date/Time 04/18/22 16:40
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CFA-1660

Date/Time 04/18/22 17:45
 Raw Sample Received by: RJ CFA-1660
 Raw Sample Relinquished by: KS (WC)