


7470A

LB119746

INSTRUMENT ID: CVL



Linear

A= 0.0000e+000

B= 2.3142e-004

C= 5.2595e-005

Rho= 0.9999826

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/6/11
0.0	0.000	0.003	0.003	14	0.000	14					-
0.2	0.200	0.203	0.003	879	0.0 %	879					2
2.5	2.500	2.483	-0.017	10731	0.0 %	10731					-1
5.0	5.000	4.990	-0.010	21564	0.0 %	21564					0
7.5	7.500	7.543	0.043	32596	0.0 %	32596					1
10.0	10.000	9.976	-0.024	43108	0.0 %	43108					0

LB119746 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Apr 2022 12:45:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	14	-	0.0000	7470A	PPB	19 Apr 2022 12:47:59	S	Std
0.2 - 1	50.2	879	-	0.2000	7470A	PPB	19 Apr 2022 12:50:16	S	Std
2.5 - 1	52.5	10731	-	2.5000	7470A	PPB	19 Apr 2022 12:52:33	S	Std
5.0 - 1	55	21564	-	5.0000	7470A	PPB	19 Apr 2022 12:54:49	S	Std
7.5 - 1	57.5	32596	-	7.5000	7470A	PPB	19 Apr 2022 12:57:05	S	Std
10.0 - 1	510	43108	-	10.0000	7470A	PPB	19 Apr 2022 12:59:21	S	Std
ICV84 - 1	ICV84	15962	3.6940	-	7470A	PPB	19 Apr 2022 13:02:12	U	SMPL
ICB84 - 1	ICB84	-82	-0.0189	-	7470A	PPB	19 Apr 2022 13:04:26	U	SMPL
CCV96 - 1	CCV96	21707	5.0235	-	7470A	PPB	19 Apr 2022 13:06:42	U	SMPL
CCB96 - 1	CCB96	-74	-0.0171	-	7470A	PPB	19 Apr 2022 13:08:56	U	SMPL
CRA - 1	CRA	833	0.1928	-	7470A	PPB	19 Apr 2022 13:11:11	U	SMPL
HighStd - 1	HighStd	42465	9.8273	-	7470A	PPB	19 Apr 2022 13:13:26	U	SMPL
ChkStd - 1	ChkStd	30645	7.0919	-	7470A	PPB	19 Apr 2022 13:15:41	U	SMPL
PB144196BL - 1	PBW	-157	-0.0363	-	7470A	PPB	19 Apr 2022 13:17:58	U	SMPL
PB144196BS - 1	LCSW	17440	4.0360	-	7470A	PPB	19 Apr 2022 13:20:14	U	SMPL
N2439-02 - 1	WARD-OIL	120	0.0278	-	7470A	PPB	19 Apr 2022 13:24:47	U	SMPL
N2439-04 - 1	WARD-OD	6	0.0014	-	7470A	PPB	19 Apr 2022 13:27:04	U	SMPL
N2442-02 - 1	124035	20	0.0047	-	7470A	PPB	19 Apr 2022 13:29:20	U	SMPL
N2442-04 - 1	124030	39	0.0091	-	7470A	PPB	19 Apr 2022 13:31:35	U	SMPL
CCV97 - 1	CCV97	21318	4.9334	-	7470A	PPB	19 Apr 2022 13:33:50	U	SMPL
CCB97 - 1	CCB97	-59	-0.0136	-	7470A	PPB	19 Apr 2022 13:36:06	U	SMPL
N2442-06 - 1	124042	26	0.0061	-	7470A	PPB	19 Apr 2022 13:38:22	U	SMPL
N2452-02 - 1	WC-15-1-2	231	0.0535	-	7470A	PPB	19 Apr 2022 13:40:38	U	SMPL
N2452-06 - 1	WC-15-1-3	78	0.0181	-	7470A	PPB	19 Apr 2022 13:42:53	U	SMPL
N2452-10 - 1	WC-15-1-4	68	0.0158	-	7470A	PPB	19 Apr 2022 13:45:09	U	SMPL
N2452-14 - 1	WC-15-1-5	38	0.0088	-	7470A	PPB	19 Apr 2022 13:47:25	U	SMPL
N2452-14DUP - 1	WC-15-1-5DUP	38	0.0088	-	7470A	PPB	19 Apr 2022 13:49:41	U	SMPL
N2452-14MS - 1	WC-15-1-5MS	17419	4.0311	-	7470A	PPB	19 Apr 2022 13:51:57	U	SMPL
N2452-14MSD - 1	WC-15-1-5MSD	16600	3.8416	-	7470A	PPB	19 Apr 2022 13:54:14	U	SMPL
PB144205BL - 1	PBW	-154	-0.0356	-	7470A	PPB	19 Apr 2022 13:56:31	U	SMPL
PB144205BS - 1	LCSW	14917	3.4521	-	7470A	PPB	19 Apr 2022 13:58:47	U	SMPL
CCV98 - 1	CCV98	21061	4.8740	-	7470A	PPB	19 Apr 2022 14:01:02	U	SMPL
CCB98 - 1	CCB98	-95	-0.0219	-	7470A	PPB	19 Apr 2022 14:03:18	U	SMPL
N2451-07 - 1	FB220415	-26	-0.0060	-	7470A	PPB	19 Apr 2022 14:05:33	U	SMPL
N2459-18 - 1	4922A	344	0.0797	-	7470A	PPB	19 Apr 2022 14:07:48	U	SMPL
N2459-20 - 1	4922B	341	0.0790	-	7470A	PPB	19 Apr 2022 14:10:04	U	SMPL
N2462-05 - 1	295-OW	86	0.0200	-	7470A	PPB	19 Apr 2022 14:12:19	U	SMPL
N2462-05DUP - 1	295-OWDUP	55	0.0128	-	7470A	PPB	19 Apr 2022 14:14:35	U	SMPL
N2462-05MS - 1	295-OWMS	16339	3.7812	-	7470A	PPB	19 Apr 2022 14:19:30	U	SMPL
N2462-05MSD - 1	295-OWMSD	16292	3.7703	-	7470A	PPB	19 Apr 2022 14:21:45	U	SMPL
CCV99 - 1	CCV99	20620	4.7719	-	7470A	PPB	19 Apr 2022 14:24:02	U	SMPL
CCB99 - 1	CCB99	-87	-0.0201	-	7470A	PPB	19 Apr 2022 14:26:19	U	SMPL
PB144163TB - 1	PB144163TB	-10	-0.0023	-	7470A	PPB	19 Apr 2022 14:28:36	U	SMPL
N2452-14LX5 - 1		-7	-0.0016	-	7470A	PPB	19 Apr 2022 14:30:51	U	SMPL
N2452-14A - 1		16564	3.8333	-	7470A	PPB	19 Apr 2022 14:33:06	U	SMPL
N2462-05LX5 - 1		-64	-0.0148	-	7470A	PPB	19 Apr 2022 14:35:22	U	SMPL
N2462-05A - 1		15704	3.6343	-	7470A	PPB	19 Apr 2022 14:37:37	U	SMPL
CCV01 - 1	CCV01	20648	4.7784	-	7470A	PPB	19 Apr 2022 14:45:25	U	SMPL
CCB01 - 1	CCB01	-73	-0.0168	-	7470A	PPB	19 Apr 2022 14:47:41	U	SMPL