



Extraction and Analytical Summary Report

Analysis Method: 9071B

Test: TPH

Run Number: LB119749

Analysis Date: 04/19/2022

BalanceID: WCSC-6

OvenID: EXT OVEN-3

ANALYST: JIGNESH

REVIEWED BY: apatel

Extraction Date: 04/19/2022

Extraction IN Time: 08:11

Extraction OUT Time: 08:35

Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Weight (g)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB119749BL	LB119749BL	SOLID		20.03	100	3.0854	3.0854	3.02	3.0855	3.0855	0.0001	4.99
2	LB119749BS	LB119749BS	SOLID		20.04	100	3.0969	3.0969	3.01	3.0988	3.0988	0.0019	94.81
3	N2394-01	CS-1- (DUMPSTER-RIGHT)	SOLID		20.02	100	3.0718	3.0718	3.03	3.1509	3.1509	0.0791	3951.05
4	N2394-03	CS-2- (DUMPSTER-RIGHT)	SOLID		20.02	100	3.0007	3.0007	3.04	3.0547	3.0547	0.0540	2697.3
5	N2394-05	CS-3-SOIL-PILE	SOLID		20.03	100	3.0912	3.0912	3.02	3.1086	3.1086	0.0174	868.7
6	N2394-07	CS-4-SOIL-PILE	SOLID		20.04	100	2.8703	2.8703	3.03	2.8891	2.8891	0.0188	938.12
7	N2394-09	CS-5- (DUMPSTER-LEFT)	SOLID		20.02	100	3.0896	3.0896	3.04	3.0936	3.0936	0.0040	199.8
8	N2394-11	CS-6- (DUMPSTER-LEFT)	SOLID		20.03	100	2.9650	2.9650	3.02	2.9664	2.9664	0.0014	69.9
9	N2394-13	CS-7-SOIL-PILE-BELOW-E	SOLID		20.04	100	2.9718	2.9718	3.03	3.0033	3.0033	0.0315	1571.86
10	N2394-15	CS-8	SOLID		20.02	100	3.0875	3.0875	3.04	3.1024	3.1024	0.0149	744.26
11	N2394-15DUP	CS-8DUP	SOLID		20.01	100	2.9130	2.9130	3.02	2.9277	2.9277	0.0147	734.63
12	N2394-15MS	CS-8	SOLID		20.03	100	2.8284	2.8284	3.01	2.8548	2.8548	0.0264	1318.02
13	N2394-15MSD	CS-8	SOLID		20.04	100	3.0126	3.0126	3.03	3.0390	3.0390	0.0264	1317.37

QC Batch# LB119749Test: TPHAnalysis Date: 04/19/2022**Chemicals Used:**

Chemical Name	Chemical Lot #
HEXANE	W2885
pH Paper 0-14	NA
Sodium Sulfate	EP2226
1:1 HCL	NA
Silica Gel	W2839
Sand	E2865

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSS	2.00 ML	WP95589
LCSSD	NA	NA
MS/MSD	2.00 ML	WP95590

BALANCE CALIBRATION / OVEN Dessicator DataAnalytical Balance ID # : WC SC-6Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 10:01
1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 09:25
Bal Check Time: 08:15 Out OVEN TEMP1: 70 °C Dessicator Time Out1 10:30
Out Time1 10:00

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 11:31
1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 11:00
Bal Check Time: 12:15 Out OVEN TEMP2: 71 °C Dessicator Time Out2 12:05
Out Time2: 11:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph n2394

WorkList ID : 158790

Department : Wet-Chemistry

Date : 04-19-2022 08:06:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/26/2022	Solid	N2394-01	TPH	Cool 4 deg C	URSC05	N12	CS-1-(DUMPSTER-RIGHT)	04/13/2022	9071B
04/26/2022	Solid	N2394-03	TPH	Cool 4 deg C	URSC05	N12	CS-2--(DUMPSTER-RIGHT)	04/13/2022	9071B
04/26/2022	Solid	N2394-05	TPH	Cool 4 deg C	URSC05	N12	CS-3-SOIL-PILE	04/13/2022	9071B
04/26/2022	Solid	N2394-07	TPH	Cool 4 deg C	URSC05	N12	CS-4-SOIL-PILE	04/13/2022	9071B
04/26/2022	Solid	N2394-09	TPH	Cool 4 deg C	URSC05	N12	CS-5-(DUMPSTER-LEFT)	04/13/2022	9071B
04/26/2022	Solid	N2394-11	TPH	Cool 4 deg C	URSC05	N12	CS-6-(DUMPSTER-LEFT)	04/13/2022	9071B
04/26/2022	Solid	N2394-13	TPH	Cool 4 deg C	URSC05	N12	CS-7-SOIL-PILE-BELOW-BRIDGE	04/13/2022	9071B
04/26/2022	Solid	N2394-15	TPH	Cool 4 deg C	URSC05	N12	CS-8	04/13/2022	9071B

Date/Time

04-19-22 08:10

Raw Sample Received by:

JP(CDC)

Raw Sample Relinquished by:

CP SM

Date/Time

04-19-22

Raw Sample Received by:

CP SM

Raw Sample Relinquished by:

JP(CDC)