


LB119752

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5105e-004 *slope*C= 3.3869e-002 *Intercept*

Rho= 0.9998865

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%I)
0.00	0.000	0.054	0.054	80	0.000	80					<u>12</u>
0.05	0.050					185	0				
0.20	0.200	0.224	0.024	759	0.0 %	759					-1
2.50	2.500	2.468	-0.032	9695	0.0 %	9695					-2
5.00	5.000	4.897	-0.103	19372	0.0 %	19372					0
7.50	7.500	7.498	-0.002	29732	0.0 %	29732					1
10.0	10.000	10.059	0.059	39931	0.0 %	39931					

LB119752
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	80	-		0 SFAM01.1	PPB	4/19/2022 15:16	Std
	0.2 S01	759	-		0.2 SFAM01.1	PPB	4/19/2022 15:18	Std
	2.5 S02	9695	-		2.5 SFAM01.1	PPB	4/19/2022 15:21	Std
	5 S03	19372	-		5 SFAM01.1	PPB	4/19/2022 15:23	Std
	7.5 S04	29732	-		7.5 SFAM01.1	PPB	4/19/2022 15:25	Std
	10 S05	39931	-		10 SFAM01.1	PPB	4/19/2022 15:28	Std
ICV085	ICV085	14257	3.6131 -		SFAM01.1	PPB	4/19/2022 15:30	SMPL
ICB085	ICB085	-54	0.0203 -		SFAM01.1	PPB	4/19/2022 15:33	SMPL
CCV002	CCV002	19542	4.9399 -		SFAM01.1	PPB	4/19/2022 15:35	SMPL
CCB002	CCB002	-54	0.0203 -		SFAM01.1	PPB	4/19/2022 15:37	SMPL
PB144225BL	PBW225	108	0.061 -		SFAM01.1	PPB	4/19/2022 15:39	SMPL
PB144192TB	LEB192	160	0.074 -		SFAM01.1	PPB	4/19/2022 15:42	SMPL
N2371-01	BH081	266	0.1006 -		SFAM01.1	PPB	4/19/2022 15:44	SMPL
N2371-02	BH082	355	0.123 -		SFAM01.1	PPB	4/19/2022 15:46	SMPL
N2371-03	BH083	193	0.0823 -		SFAM01.1	PPB	4/19/2022 15:48	SMPL
N2371-04	BH084	190	0.0816 -		SFAM01.1	PPB	4/19/2022 15:51	SMPL
N2371-05	BH085	99	0.0587 -		SFAM01.1	PPB	4/19/2022 15:53	SMPL
N2371-06	BH086	27	0.0406 -		SFAM01.1	PPB	4/19/2022 15:55	SMPL
N2371-07	BH087	155	0.0728 -		SFAM01.1	PPB	4/19/2022 15:58	SMPL
N2371-08	BH088	461	0.1496 -		SFAM01.1	PPB	4/19/2022 16:00	SMPL
N2371-09	BH089	139	0.0688 -		SFAM01.1	PPB	4/19/2022 16:02	SMPL
N2371-10	BH090	26	0.0404 -		SFAM01.1	PPB	4/19/2022 16:04	SMPL
N2371-11	BH091	16	0.0379 -		SFAM01.1	PPB	4/19/2022 16:07	SMPL
N2371-12	BH092	31	0.0417 -		SFAM01.1	PPB	4/19/2022 16:09	SMPL
N2371-13	BH093	25	0.0401 -		SFAM01.1	PPB	4/19/2022 16:11	SMPL
N2371-14	BH094	98	0.0585 -		SFAM01.1	PPB	4/19/2022 16:13	SMPL
N2371-15	BH095	407	0.136 -		SFAM01.1	PPB	4/19/2022 16:16	SMPL
CCV003	CCV003	18469	4.6705 -		SFAM01.1	PPB	4/19/2022 16:18	SMPL
CCB003	CCB003	-21	0.0286 -		SFAM01.1	PPB	4/19/2022 16:20	SMPL