



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119758

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	04/19/2022	17:45
2	CAL2	1	Water	NA	NA	20.4	7.01	04/19/2022	17:46
3	CAL3	1	Water	NA	NA	20.4	10.03	04/19/2022	17:47
4	ICV	1	Water	NA	NA	20.4	6.98	04/19/2022	17:48
5	CCV1	1	Water	NA	NA	20.3	2.01	04/19/2022	17:49
6	N2474-02	1	Solid	20.02	20	20.7	12.61	04/19/2022	18:00
7	N2474-02DUP	1	Solid	20.01	20	20.8	12.62	04/19/2022	18:01
8	N2474-05	1	Solid	20.03	20	20.4	12.65	04/19/2022	18:03
9	N2480-04	1	Solid	20.04	20	20.2	8.68	04/19/2022	18:05
10	N2480-08	1	Solid	20.03	20	20.6	8.73	04/19/2022	18:10
11	N2480-12	1	Solid	20.04	20	20.7	7.15	04/19/2022	18:15
12	N2480-16	1	Solid	20.03	20	20.3	8.42	04/19/2022	18:16
13	CCV2	1	Water	NA	NA	20.5	12.03	04/19/2022	18:20

WB 119758

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIVITY N2474 WorkList ID: 158817 Department: Wet-Chemistry Date: 04-19-2022 14:26:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
04/19/2022	Solid	N2480-04	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-16-1A-1	04/19/2022	9045D
04/19/2022	Solid	N2480-08	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-16-1A-2	04/19/2022	9045D
04/19/2022	Solid	N2480-12	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-16-1A-3	04/19/2022	9045D
04/19/2022	Solid	N2480-16	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-16-1A-4	04/19/2022	9045D
04/19/2022	Solid	N2474-02	Corrosivity	Cool 4 deg C	TULL02	P11	TMS-JET-GROUT-SPOILS-A	04/19/2022	9045D
04/19/2022	Solid	N2474-05	Corrosivity	Cool 4 deg C	TULL02	P11	TMS-JET-GROUT-SPOILS-B	04/19/2022	9045D

Date/Time 04.19.22 17:40
 Raw Sample Received by: CP 5M
 Raw Sample Relinquished by: CP 5M

Date/Time 04.19.22 18:30
 Raw Sample Received by: CP 5M
 Raw Sample Relinquished by: JP 100C