

LB 119760


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

52334



A= 0.0000e+000

B= 1.9133e-004

C= -2.1074e-002

Rho= 0.9999722

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1))
0.0	0.000	-0.015	-0.015	34	0.000	34					
0.2	0.200	0.184	-0.016	1073	0.0 %	1073					-8
2.5	2.500	2.515	0.015	13255	0.0 %	13255					1
5.0	5.000	5.054	0.054	26523	0.0 %	26523					1
7.5	7.500	7.470	-0.030	39151	0.0 %	39151					0
10.0	10.000	9.992	-0.008	52334	0.0 %	52334					0

LB119760 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 20 Apr 2022 12:15:56				
Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	20 Apr 2022 12:17:59	S Std
0.2 - 1	50.2	1073	-	0.2000	7471B	PPB	20 Apr 2022 12:20:15	S Std
2.5 - 1	52.1	13255	-	2.5000	7471B	PPB	20 Apr 2022 12:22:32	S Std
5.0 - 1	55	26523	-	5.0000	7471B	PPB	20 Apr 2022 12:24:48	S Std
7.5 - 1	57.5	39151	-	7.5000	7471B	PPB	20 Apr 2022 12:27:05	S Std
10.0 - 1	518	52334	-	10.0000	7471B	PPB	20 Apr 2022 12:29:22	S Std
ICV86 - 1	ICV86	19063	3.6263	-	7471B	PPB	20 Apr 2022 12:32:01	U SMPL
ICB86 - 1	ICB86	-75	-0.0354	-	7471B	PPB	20 Apr 2022 12:34:16	U SMPL
CCV05 - 1	CCV05	26388	5.0278	-	7471B	PPB	20 Apr 2022 12:36:31	U SMPL
CCB05 - 1	CCB05	-99	-0.0400	-	7471B	PPB	20 Apr 2022 12:38:45	U SMPL
CRA - 1	CRA	1020	0.1741	-	7471B	PPB	20 Apr 2022 12:41:02	U SMPL
HighStd - 1	HighStd	51161	9.7676	-	7471B	PPB	20 Apr 2022 12:43:17	U SMPL
ChkStd - 1	ChkStd	36822	7.0241	-	7471B	PPB	20 Apr 2022 12:45:32	U SMPL
PB144248BL - 1	PBS	226	0.0222	-	7471B	PPB	20 Apr 2022 12:50:23	U SMPL
PB144248BS - 1	LCSS	20996	3.9961	-	7471B	PPB	20 Apr 2022 12:52:39	U SMPL
N2459-01 - 1	341	270	0.0306	-	7471B	PPB	20 Apr 2022 12:54:55	U SMPL
N2459-02 - 1	201	2544	0.4657	-	7471B	PPB	20 Apr 2022 12:57:12	U SMPL
N2464-01 - 1	OK-03-041822	2395	0.4372	-	7471B	PPB	20 Apr 2022 12:59:28	U SMPL
N2466-03 - 1	E14ST-EB2-041822-0-5	4009	0.7460	-	7471B	PPB	20 Apr 2022 13:01:45	U SMPL
N2466-04 - 1	E14ST-EB2-041822-0-5-DUP	5241	0.9817	-	7471B	PPB	20 Apr 2022 13:04:00	U SMPL
CCV06 - 1	CCV06	25529	4.8634	-	7471B	PPB	20 Apr 2022 13:06:16	U SMPL
CCB06 - 1	CCB06	-56	-0.0318	-	7471B	PPB	20 Apr 2022 13:08:30	U SMPL
N2466-05 - 1	E14ST-EB2-041822-5-10	8199	1.5476	-	7471B	PPB	20 Apr 2022 13:10:46	U SMPL
N2476-01 - 1	PAT-01-041922	4300	0.8016	-	7471B	PPB	20 Apr 2022 13:13:00	U SMPL
N2476-03 - 1	PAT-02-041922	6699	1.2606	-	7471B	PPB	20 Apr 2022 13:15:16	U SMPL
N2476-05 - 1	PAT-03-041922	6185	1.1623	-	7471B	PPB	20 Apr 2022 13:17:31	U SMPL
N2477-01 - 1	371	1118	0.1928	-	7471B	PPB	20 Apr 2022 13:19:47	U SMPL
N2478-01 - 1	279	1440	0.2544	-	7471B	PPB	20 Apr 2022 13:22:03	U SMPL
N2479-04 - 1	E14ST-EB4-041922-0-5	5669	1.0636	-	7471B	PPB	20 Apr 2022 13:33:52	U SMPL
N2479-04DUP - 1	E14ST-EB4-041922-0-5DUP	5459	1.0234	-	7471B	PPB	20 Apr 2022 13:36:09	U SMPL
N2479-05 - 1	N2179-04MS	28578	5.4468	-	7471B	PPB	20 Apr 2022 13:38:25	U SMPL
N2479-06 - 1	N2179-04MSD	30220	5.7609	-	7471B	PPB	20 Apr 2022 13:40:41	U SMPL
CCV07 - 1	CCV07	28743	5.4783	-	7471B	PPB	20 Apr 2022 13:42:57	U SMPL
CCB07 - 1	CCB07	-65	-0.0335	-	7471B	PPB	20 Apr 2022 13:45:14	U SMPL
N2479-08 - 1	E14ST-EB4-041922-5-10	58963	11.2604	-	7471B	PPB	20 Apr 2022 13:47:30	U SMPL
N2479-04LX5 - 1		878	0.1469	-	7471B	PPB	20 Apr 2022 13:49:46	U SMPL
N2479-04A - 1		28440	5.4204	-	7471B	PPB	20 Apr 2022 13:52:04	U SMPL
N2479-08DLX2 - 1		30336	5.7831	-	7471B	PPB	20 Apr 2022 13:54:19	U SMPL
CCV08 - 1	CCV08	27386	5.2187	-	7471B	PPB	20 Apr 2022 13:59:06	U SMPL
CCB08 - 1	CCB08	-52	-0.0310	-	7471B	PPB	20 Apr 2022 14:01:23	U SMPL