


LB119765

SFAM01.1



Linear

A= 0.0000e+000

B= 2.2316e-004 slope

C= -1.6205e-001 y-intercept

Rho= 0.9998958

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	-0.079	-0.079	374	0.000	374					1
0.05	0.050					185	0				25
0.20	0.200	0.249	0.049	1843	0.0 %	1843					0
2.50	2.500	2.488	-0.012	11873	0.0 %	11873					0
5.00	5.000	5.078	0.078	23483	0.0 %	23483					2
7.50	7.500	7.509	0.009	34373	0.0 %	34373					0
10.0	10.000	9.955	-0.045	45334	0.0 %	45334					0

LB119765
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	374	-		0 SFAM01.1	PPB	4/20/2022 14:59	Std
	0.2 S01	1843	-		0.2 SFAM01.1	PPB	4/20/2022 15:01	Std
	2.5 S02	11873	-		2.5 SFAM01.1	PPB	4/20/2022 15:04	Std
	5 S03	23483	-		5 SFAM01.1	PPB	4/20/2022 15:06	Std
	7.5 S04	34373	-		7.5 SFAM01.1	PPB	4/20/2022 15:08	Std
	10 S05	45334	-		10 SFAM01.1	PPB	4/20/2022 15:10	Std
ICV087	ICV087	18114	3.8803 -		SFAM01.1	PPB	4/20/2022 15:13	SMPL
ICB087	ICB087	611	-0.0257 -		SFAM01.1	PPB	4/20/2022 15:16	SMPL
CCV009	CCV009	23203	5.0159 -		SFAM01.1	PPB	4/20/2022 15:18	SMPL
CCB009	CCB009	608	-0.0264 -		SFAM01.1	PPB	4/20/2022 15:20	SMPL
PB144245BL	PBW245	685	-0.0092 -		SFAM01.1	PPB	4/20/2022 15:22	SMPL
N2346-02	MDBSQ6	19517	4.1934 -		SFAM01.1	PPB	4/20/2022 15:24	SMPL
N2346-03	MDBSQ8	1606	0.1963 -		SFAM01.1	PPB	4/20/2022 15:27	SMPL
N2346-04	MDBSQ9	414	-0.0697 -		SFAM01.1	PPB	4/20/2022 15:29	SMPL
N2346-05	MDBSQ9D	487	-0.0534 -		SFAM01.1	PPB	4/20/2022 15:31	SMPL
N2346-06	MDBSQ9S	5014	0.9569 -		SFAM01.1	PPB	4/20/2022 15:34	SMPL
N2346-07	MDBSR0	468	-0.0576 -		SFAM01.1	PPB	4/20/2022 15:36	SMPL
N2346-08	MDBSR1	13402	2.8288 -		SFAM01.1	PPB	4/20/2022 15:38	SMPL
N2346-09	MDBSR2	1464	0.1647 -		SFAM01.1	PPB	4/20/2022 15:40	SMPL
N2346-10	MDBSR3	2031	0.2912 -		SFAM01.1	PPB	4/20/2022 15:43	SMPL
N2346-11	MDBSR4	598	-0.0286 -		SFAM01.1	PPB	4/20/2022 15:45	SMPL
N2346-12	MDBSR5	401	-0.0726 -		SFAM01.1	PPB	4/20/2022 15:47	SMPL
PB144246BL	PBW246	228	-0.1112 -		SFAM01.1	PPB	4/20/2022 15:49	SMPL
N2347-01	MDBST4	790	0.0142 -		SFAM01.1	PPB	4/20/2022 15:52	SMPL
N2347-02	MDBST5	218	-0.1134 -		SFAM01.1	PPB	4/20/2022 15:54	SMPL
N2347-03	MDBST6	210	-0.1152 -		SFAM01.1	PPB	4/20/2022 15:56	SMPL
N2347-04	MDBST7	285	-0.0984 -		SFAM01.1	PPB	4/20/2022 15:58	SMPL
N2347-05	MDBST8	278	-0.1 -		SFAM01.1	PPB	4/20/2022 16:01	SMPL
N2347-06	MDBST9	250	-0.1063 -		SFAM01.1	PPB	4/20/2022 16:03	SMPL
N2347-07	MDBSU0	244	-0.1076 -		SFAM01.1	PPB	4/20/2022 16:05	SMPL
CCV010	CCV010	22923	4.9535 -		SFAM01.1	PPB	4/20/2022 16:08	SMPL
CCB010	CCB010	638	-0.0197 -		SFAM01.1	PPB	4/20/2022 16:10	SMPL
N2347-08	MDBSU0D	215	-0.1141 -		SFAM01.1	PPB	4/20/2022 16:12	SMPL
N2347-09	MDBSU0S	4745	0.8969 -		SFAM01.1	PPB	4/20/2022 16:14	SMPL
N2347-10	MDBSU1	440	-0.0639 -		SFAM01.1	PPB	4/20/2022 16:17	SMPL
N2347-11	MDBSU2	379	-0.0775 -		SFAM01.1	PPB	4/20/2022 16:19	SMPL
N2347-12	MDBSU3	277	-0.1002 -		SFAM01.1	PPB	4/20/2022 16:21	SMPL
N2347-13	MDBSU5	399	-0.073 -		SFAM01.1	PPB	4/20/2022 16:23	SMPL
CCV011	CCV011	22855	4.9383 -		SFAM01.1	PPB	4/20/2022 16:32	SMPL
CCB011	CCB011	625	-0.0226 -		SFAM01.1	PPB	4/20/2022 16:35	SMPL