



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119767

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/20/2022	17:20
2	CAL2	1	Water	NA	NA	20.3	7.00	04/20/2022	17:21
3	CAL3	1	Water	NA	NA	20.3	10.03	04/20/2022	17:22
4	ICV	1	Water	NA	NA	20.4	6.98	04/20/2022	17:25
5	CCV1	1	Water	NA	NA	20.4	2.01	04/20/2022	17:26
6	N2511-06	1	Water	NA	NA	20.1	6.38	04/20/2022	17:29
7	N2511-06DUP	1	Water	NA	NA	20.2	6.39	04/20/2022	17:30
8	N2511-07	1	Water	NA	NA	20.2	5.25	04/20/2022	17:32
9	N2511-08	1	Water	NA	NA	20.9	8.10	04/20/2022	17:35
10	N2511-09	1	Water	NA	NA	20.6	5.95	04/20/2022	17:36
11	N2511-10	1	Water	NA	NA	20.7	7.14	04/20/2022	17:40
12	N2511-11	1	Water	NA	NA	20.2	6.99	04/20/2022	17:42
13	N2511-12	1	Water	NA	NA	20.1	6.10	04/20/2022	17:45
14	N2511-13	1	Water	NA	NA	20.2	6.04	04/20/2022	17:48
15	CCV2	1	Water	NA	NA	20.6	12.03	04/20/2022	17:49

18 114767

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2511 WorkList ID : 158852 Department : Wet-Chemistry Date : 04-20-2022 17:11:21

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Raw Sample Storage Location		
04/20/2022	Water	N2511-06	pH <u>C</u>	Cool 4 deg C	PSEG03	P11 1530	04/20/2022	9040C
04/20/2022	Water	N2511-07	pH <u>C</u>	Cool 4 deg C	PSEG03	P11 1502	04/20/2022	9040C
04/20/2022	Water	N2511-08	pH <u>C</u>	Cool 4 deg C	PSEG03	P11 1551	04/20/2022	9040C
04/20/2022	Water	N2511-09	pH <u>A</u>	Cool 4 deg C	PSEG03	P11 1590	04/20/2022	9040C
04/20/2022	Water	N2511-10	pH <u>C</u>	Cool 4 deg C	PSEG03	P11 1593	04/20/2022	9040C
04/20/2022	Water	N2511-11	pH <u>C</u>	Cool 4 deg C	PSEG03	P11 1592	04/20/2022	9040C
04/20/2022	Water	N2511-12	pH <u>C</u>	Cool 4 deg C	PSEG03	P11 1591	04/20/2022	9040C
04/20/2022	Water	N2511-13	pH <u>C</u>	Cool 4 deg C	PSEG03	P11 1553	04/20/2022	9040C

Date/Time 04/20/22 17:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 04/20/22 18:20
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)