


LB119771

7470A

INSTRUMENT ID: CV1

 μ Abs.:

43602



Linear

A= 0.0000e+000

B= 2.3173e-004

C= -3.6806e-003

Rho= 0.9989381

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b)
0.0	0.000	0.018	0.018	94	0.000	94					-
0.2	0.200	0.207	0.007	909	0.0 %	909					3
2.5	2.500	2.419	-0.081	10456	0.0 %	10456					-3
5.0	5.000	5.256	0.256	22699	0.0 %	22699					5
7.5	7.500	7.199	-0.301	31081	0.0 %	31081					-4
10.0	10.000	10.100	0.100	43602	0.0 %	43602					1

LB119771 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Apr 2022 09:00:10

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	94	-	0.00007470A	PPB	21 Apr 2022 09:04:46	S	Std
0.2 - 1	50.2	909	-	0.20007470A	PPB	21 Apr 2022 09:07:03	S	Std
2.5 - 1	52.5	10456	-	2.50007470A	PPB	21 Apr 2022 09:09:19	S	Std
5.0 - 1	55	22699	-	5.00007470A	PPB	21 Apr 2022 09:11:36	S	Std
7.5 - 1	57.5	31081	-	7.50007470A	PPB	21 Apr 2022 09:13:52	S	Std
10.0 - 1	510	43602	-	10.00007470A	PPB	21 Apr 2022 09:21:18	S	Std
ICV88 - 1	ICV88	17094	3.9576	-7470A	PPB	21 Apr 2022 09:26:44	U	SMPL
ICB88 - 1	ICB88	-18	-0.0079	-7470A	PPB	21 Apr 2022 09:28:59	U	SMPL
CCV12 - 1	CCV12	19938	4.6166	-7470A	PPB	21 Apr 2022 09:31:14	U	SMPL
CCB12 - 1	CCB12	0	-0.0037	-7470A	PPB	21 Apr 2022 09:33:29	U	SMPL
CRA - 1	CRA	820	0.1863	-7470A	PPB	21 Apr 2022 09:35:44	U	SMPL
HighStd - 1	HighStd	43007	9.9625	-7470A	PPB	21 Apr 2022 09:37:59	U	SMPL
ChkStd - 1	ChkStd	28994	6.7152	-7470A	PPB	21 Apr 2022 09:40:14	U	SMPL
PB144255BL - 1	PBW	-157	-0.0401	-7470A	PPB	21 Apr 2022 09:42:31	U	SMPL
PB144255BS - 1	LCSW	15209	3.5208	-7470A	PPB	21 Apr 2022 09:47:19	U	SMPL
N2468-04 - 1	WC-15-1-6	236	0.0510	-7470A	PPB	21 Apr 2022 09:49:35	U	SMPL
N2468-08 - 1	WC-15-1-7	122	0.0246	-7470A	PPB	21 Apr 2022 09:51:52	U	SMPL
N2468-12 - 1	WC-15-1-8	100	0.0195	-7470A	PPB	21 Apr 2022 09:54:09	U	SMPL
N2474-02 - 1	TMS-JET-GROUT-SPOILS-A	73	0.0132	-7470A	PPB	21 Apr 2022 09:56:26	U	SMPL
N2474-05 - 1	TMS-JET-GROUT-SPOILS-B	89	0.0169	-7470A	PPB	21 Apr 2022 09:58:41	U	SMPL
CCV13 - 1	CCV13	19819	4.5890	-7470A	PPB	21 Apr 2022 10:00:56	U	SMPL
CCB13 - 1	CCB13	35	0.0044	-7470A	PPB	21 Apr 2022 10:03:11	U	SMPL
N2480-04 - 1	WC-16-1A-1	71	0.0128	-7470A	PPB	21 Apr 2022 10:05:26	U	SMPL
N2480-08 - 1	WC-16-1A-2	180	0.0380	-7470A	PPB	21 Apr 2022 10:07:41	U	SMPL
N2480-12 - 1	WC-16-1A-3	136	0.0278	-7470A	PPB	21 Apr 2022 10:09:56	U	SMPL
N2480-16 - 1	WC-16-1A-4	152	0.0315	-7470A	PPB	21 Apr 2022 10:12:12	U	SMPL
N2480-16DUP - 1	WC-16-1A-4DUP	363	0.0804	-7470A	PPB	21 Apr 2022 10:14:27	U	SMPL
N2480-16MS - 1	WC-16-1A-4MS	16242	3.7601	-7470A	PPB	21 Apr 2022 10:16:43	U	SMPL
N2480-16MSD - 1	WC-16-1A-4MSD	16243	3.7604	-7470A	PPB	21 Apr 2022 10:19:00	U	SMPL
PB144226TB - 1	PB144226TB	1	-0.0034	-7470A	PPB	21 Apr 2022 10:21:17	U	SMPL
N2480-16LX5 - 1		-18	-0.0079	-7470A	PPB	21 Apr 2022 10:23:34	U	SMPL
N2480-16A - 1		17903	4.1450	-7470A	PPB	21 Apr 2022 10:25:49	U	SMPL
CCV14 - 1	CCV14	19778	4.5795	-7470A	PPB	21 Apr 2022 10:28:04	U	SMPL
CCB14 - 1	CCB14	19	0.0007	-7470A	PPB	21 Apr 2022 10:30:20	U	SMPL