

LB119775


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55330



A= 0.0000e+000

B= 1.8024e-004

C= -2.9181e-002

Rho= 0.9999330

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.011	-0.011	101	0.000	101					-
0.2	0.200	0.190	-0.010	1215	0.0 %	1215					-5
2.5	2.500	2.469	-0.031	13861	0.0 %	13861					-1
5.0	5.000	5.072	0.072	28304	0.0 %	28304					1
7.5	7.500	7.537	0.037	41977	0.0 %	41977					0
10.0	10.000	9.943	-0.057	55330	0.0 %	55330					-1

LB119775 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Apr 2022 11:15:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	101	-	0.00007471B	PPB		21 Apr 2022 11:17:48	S	Std
0.2 - 1	50.2	1215	-	0.20007471B	PPB		21 Apr 2022 11:20:05	S	Std
2.5 - 1	52.5	13861	-	2.50007471B	PPB		21 Apr 2022 11:22:22	S	Std
5.0 - 1	55	28304	-	5.00007471B	PPB		21 Apr 2022 11:24:39	S	Std
7.5 - 1	57.5	41977	-	7.50007471B	PPB		21 Apr 2022 11:26:55	S	Std
10.0 - 1	510	55330	-	10.00007471B	PPB		21 Apr 2022 11:29:11	S	Std
ICV89 - 1	ICV89	22123	3.9582	-7471B	PPB		21 Apr 2022 11:31:54	U	SMPL
ICB89 - 1	ICB89	0	-0.0292	-7471B	PPB		21 Apr 2022 11:34:10	U	SMPL
CCV15 - 1	CCV15	28122	5.0394	-7471B	PPB		21 Apr 2022 11:36:25	U	SMPL
CCB15 - 1	CCB15	1	-0.0290	-7471B	PPB		21 Apr 2022 11:38:40	U	SMPL
CRA - 1	CRA	1186	0.1846	-7471B	PPB		21 Apr 2022 11:40:55	U	SMPL
HighStd - 1	HighStd	56386	10.1336	-7471B	PPB		21 Apr 2022 11:43:10	U	SMPL
ChkStd - 1	ChkStd	38565	6.9216	-7471B	PPB		21 Apr 2022 11:45:25	U	SMPL
PB144274BL - 1	PBS	-77	-0.0431	-7471B	PPB		21 Apr 2022 11:47:42	U	SMPL
PB144274BS - 1	LCSS	23371	4.1831	-7471B	PPB		21 Apr 2022 11:49:59	U	SMPL
N2431-01 - 1	WC-15-1-1	648	0.0876	-7471B	PPB		21 Apr 2022 11:52:15	U	SMPL
N2452-01 - 1	WC-15-1-2	14167	2.5242	-7471B	PPB		21 Apr 2022 11:54:33	U	SMPL
N2452-05 - 1	WC-15-1-3	5968	1.0465	-7471B	PPB		21 Apr 2022 11:56:49	U	SMPL
N2452-09 - 1	WC-15-1-4	31564	5.6598	-7471B	PPB		21 Apr 2022 11:59:06	U	SMPL
N2452-13 - 1	WC-15-1-5	11058	1.9639	-7471B	PPB		21 Apr 2022 12:01:22	U	SMPL
CCV16 - 1	CCV16	28740	5.1508	-7471B	PPB		21 Apr 2022 12:03:37	U	SMPL
CCB16 - 1	CCB16	7	-0.0279	-7471B	PPB		21 Apr 2022 12:05:52	U	SMPL
N2468-03 - 1	WC-15-1-6	23324	4.1746	-7471B	PPB		21 Apr 2022 12:08:09	U	SMPL
N2468-07 - 1	WC-15-1-7	18490	3.3034	-7471B	PPB		21 Apr 2022 12:10:24	U	SMPL
N2468-11 - 1	WC-15-1-8	17797	3.1785	-7471B	PPB		21 Apr 2022 12:11:39	U	SMPL
N2480-03 - 1	WC-16-1A-1	15920	2.8402	-7471B	PPB		21 Apr 2022 12:13:54	U	SMPL
N2480-07 - 1	WC-16-1A-2	19551	3.4946	-7471B	PPB		21 Apr 2022 12:16:10	U	SMPL
N2480-11 - 1	WC-16-1A-3	20024	3.5799	-7471B	PPB		21 Apr 2022 12:18:26	U	SMPL
N2480-15 - 1	WC-16-1A-4	12793	2.2766	-7471B	PPB		21 Apr 2022 12:20:43	U	SMPL
N2491-01 - 1	RBR-200046-333	614	0.0815	-7471B	PPB		21 Apr 2022 12:23:00	U	SMPL
N2493-01 - 1	72-11944	7135	1.2568	-7471B	PPB		21 Apr 2022 12:25:18	U	SMPL
N2494-01 - 1	MH-1S	1000	0.1511	-7471B	PPB		21 Apr 2022 12:27:33	U	SMPL
CCV17 - 1	CCV17	28382	5.0863	-7471B	PPB		21 Apr 2022 12:29:49	U	SMPL
CCB17 - 1	CCB17	-4	-0.0299	-7471B	PPB		21 Apr 2022 12:32:04	U	SMPL
N2494-01DUP - 1	MH-1SDUP	1085	0.1664	-7471B	PPB		21 Apr 2022 12:34:20	U	SMPL
N2494-01MS - 1	MH-1SMS	23222	4.1563	-7471B	PPB		21 Apr 2022 12:36:35	U	SMPL
N2494-01MSD - 1	MH-1MSD	22007	3.9373	-7471B	PPB		21 Apr 2022 12:38:51	U	SMPL
N2494-04 - 1	MH-1D	635	0.0853	-7471B	PPB		21 Apr 2022 12:41:07	U	SMPL
N2507-01 - 1	PL-01-042022	2892	0.4921	-7471B	PPB		21 Apr 2022 12:43:22	U	SMPL
N2494-01LX5 - 1		256	0.0170	-7471B	PPB		21 Apr 2022 12:45:38	U	SMPL
N2494-01A - 1		23415	4.1910	-7471B	PPB		21 Apr 2022 12:47:55	U	SMPL
CCV18 - 1	CCV18	28467	5.1016	-7471B	PPB		21 Apr 2022 12:50:12	U	SMPL
CCB18 - 1	CCB18	14	-0.0267	-7471B	PPB		21 Apr 2022 12:52:30	U	SMPL