

LB117783

Result archive

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : R Instrument ID : Konelab

4/21/2022 14:09

Sample id	Test id	Result	Status	Date	Errors
ICV1	Total CN	95.265	man acc	4/21/2022	
ICB1	Total CN	0.782	man acc	4/21/2022	
CCV1	Total CN	246.560	man acc	4/21/2022	
CCB1	Total CN	0.433	man acc	4/21/2022	
PB144269BL	Total CN	0.342	man acc	4/21/2022	
N2474-02	Total CN	0.384	man acc	4/21/2022	
N2474-02DUP	Total CN	0.218	man acc	4/21/2022	
N2474-05	Total CN	0.265	man acc	4/21/2022	
N2480-04	Total CN	0.448	man acc	4/21/2022	
N2480-08	Total CN	0.448	man acc	4/21/2022	
N2480-12	Total CN	0.406	man acc	4/21/2022	
N2480-16	Total CN	0.487	man acc	4/21/2022	
N2494-02	Total CN	0.412	man acc	4/21/2022	
N2494-05	Total CN	0.686	man acc	4/21/2022	
CCV2	Total CN	243.059	man acc	4/21/2022	
CCB2	Total CN	0.643	man acc	4/21/2022	
N2514-04	Total CN	0.191	man acc	4/21/2022	
N2514-08	Total CN	0.603	man acc	4/21/2022	

=====
Result archive

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

4/21/2022 14:10

Sample id	Test id	^{mpil} Result	Status	Date	Errors
CCV3	Total CN	243.435	man acc	4/21/2022	
CCB3	Total CN	1.438	man acc	4/21/2022	

Aquakem v. 7.2AQ1

Results from time period:

Thu Apr 21 13:45:37 2022

Thu Apr 21 14:04:08 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-1.1626	µg/l	4/21/2022 10:37:01	
5.OPPBCN	A	Total CN	P	4.9154	µg/l	4/21/2022 10:37:02	
10PPBCN	A	Total CN	P	9.7185	µg/l	4/21/2022 10:37:03	
50PPBCN	A	Total CN	P	50.2418	µg/l	4/21/2022 10:37:04	
100PPBCN	A	Total CN	P	100.7874	µg/l	4/21/2022 10:37:05	
250PPBCN	A	Total CN	P	251.3494	µg/l	4/21/2022 10:37:06	
500PPBCN	A	Total CN	P	499.1501	µg/l	4/21/2022 10:37:07	
ICV1	S	Total CN	P	95.2652	µg/l	4/21/2022 13:45:38	
ICB1	S	Total CN	P	0.7819	µg/l	4/21/2022 13:45:39	
CCV1	S	Total CN	P	246.5596	µg/l	4/21/2022 13:45:41	
CCB1	S	Total CN	P	0.4331	µg/l	4/21/2022 13:45:44	
PB144269BL	S	Total CN	P	0.3416	µg/l	4/21/2022 13:45:45	
N2474-02	S	Total CN	P	0.3835	µg/l	4/21/2022 13:45:47	
N2474-02DUP	S	Total CN	P	0.2184	µg/l	4/21/2022 13:53:09	
N2474-05	S	Total CN	P	0.2646	µg/l	4/21/2022 13:53:10	
N2480-04	S	Total CN	P	0.4483	µg/l	4/21/2022 13:53:11	
N2480-08	S	Total CN	P	0.4477	µg/l	4/21/2022 13:53:12	
N2480-12	S	Total CN	P	0.4059	µg/l	4/21/2022 13:53:13	
N2480-16	S	Total CN	P	0.4873	µg/l	4/21/2022 13:53:14	
N2494-02	S	Total CN	P	0.4124	µg/l	4/21/2022 13:53:15	
N2494-05	S	Total CN	P	0.6863	µg/l	4/21/2022 13:53:16	
CCV2	S	Total CN	P	243.0591	µg/l	4/21/2022 13:53:18	
CCB2	S	Total CN	P	0.6431	µg/l	4/21/2022 13:53:19	
N2514-04	S	Total CN	P	0.191	µg/l	4/21/2022 14:00:44	
N2514-08	S	Total CN	P	0.6034	µg/l	4/21/2022 14:00:45	
CCV3	S	Total CN	P	243.4347	µg/l	4/21/2022 14:00:47	
CCB3	S	Total CN	P	1.4378	µg/l	4/21/2022 14:00:48	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

4/21/2022 10:38


Test Total CN

Accepted 4/21/2022 10:38

Factor 1151
Bias 0.001

Coeff. of det. 0.999977

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	-0.000	-1.1626	0.0000	-
2	5.0PPBCN	0.005	4.9154	5.0000	-1.7
3	10PPBCN	0.009	9.7185	10.0000	-2.8
4	50PPBCN	0.044	50.2418	50.0000	0.5
5	100PPBCN	0.088	100.7874	100.0000	0.8
6	250PPBCN	0.219	251.3494	250.0000	0.5
7	500PPBCN	0.434	499.1501	500.0000	-0.2

12
4/21/22