



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119784

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/21/2022	15:25
2	CAL2	1	Water	NA	NA	20.8	7.00	04/21/2022	15:26
3	CAL3	1	Water	NA	NA	20.7	10.02	04/21/2022	15:27
4	ICV	1	Water	NA	NA	20.7	6.99	04/21/2022	15:30
5	CCV1	1	Water	NA	NA	20.6	2.01	04/21/2022	15:32
6	N2522-01	1	Water	NA	NA	20.6	5.80	04/21/2022	15:35
7	N2522-02	1	Water	NA	NA	20.3	6.51	04/21/2022	15:37
8	N2522-03	1	Water	NA	NA	20.1	7.02	04/21/2022	15:39
9	N2524-01	1	Water	NA	NA	20.1	6.30	04/21/2022	15:40
10	N2524-03	1	Water	NA	NA	20.2	6.47	04/21/2022	15:42
11	N2524-05	1	Water	NA	NA	20.3	6.17	04/21/2022	15:45
12	N2524-07	1	Water	NA	NA	20.6	6.24	04/21/2022	15:47
13	N2524-09	1	Water	NA	NA	20.5	6.27	04/21/2022	15:48
14	N2524-09DUP	1	Water	NA	NA	20.6	6.28	04/21/2022	15:49
15	CCV2	1	Water	NA	NA	20.7	12.03	04/21/2022	15:50

WB 119484

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2524 WorkList ID : 158895 Department : Wet-Chemistry Date : 04-21-2022 15:18:16

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
04/21/2022	Water	N2522-01	pH ✓	Cool 4 deg C	PSEG03	P11	NWB-1756	04/21/2022	9040C
04/21/2022	Water	N2522-02	pH ✓	Cool 4 deg C	PSEG03	P11	NWB-1757	04/21/2022	9040C
04/21/2022	Water	N2522-03	pH ✓	Cool 4 deg C	PSEG03	P11	NWB-1766	04/21/2022	9040C
04/21/2022	Water	N2524-01	pH ✓	Cool 4 deg C	PSEG03	P11	33659	04/21/2022	9040C
04/21/2022	Water	N2524-03	pH ✓	Cool 4 deg C	PSEG03	P11	33655	04/21/2022	9040C
04/21/2022	Water	N2524-05	pH ✓	Cool 4 deg C	PSEG03	P11	33656-57	04/21/2022	9040C
04/21/2022	Water	N2524-07	pH ✓	Cool 4 deg C	PSEG03	P11	33658	04/21/2022	9040C
04/21/2022	Water	N2524-09	pH ✓	Cool 4 deg C	PSEG03	P11	33660	04/21/2022	9040C

15:20

Date/Time 04-21-22 15:20
Raw Sample Received by: J.P. W.C.
Raw Sample Relinquished by: C.P. SM

16:35

Date/Time 04-21-22
Raw Sample Received by: C.P. SM
Raw Sample Relinquished by: J.P. W.C.