



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB119787

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	04/21/2022	15:05
2	N2511-06		1	12	100.00	>212.0	>212.0	04/21/2022	15:25
3	N2511-06DUP		1	12	100.00	>212.0	>212.0	04/21/2022	15:40
4	N2511-07		1	15	100.00	>212.0	>212.0	04/21/2022	15:58
5	N2511-08		1	12	100.00	>212.0	>212.0	04/21/2022	16:20
6	N2511-09		1	10	100.00	>212.0	>212.0	04/21/2022	16:45
7	N2511-10		1	14	100.00	>212.0	>212.0	04/21/2022	17:05
8	N2511-11		1	15	100.00	>212.0	>212.0	04/21/2022	17:20
9	N2511-12		1	12	100.00	>212.0	>212.0	04/21/2022	17:40
10	N2511-13		1	15	100.00	>212.0	>212.0	04/21/2022	17:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB119787

WorkList Name : FP42122 WorkList ID : 158874 Department : Wet-Chemistry Date : 04-21-2022 11:48:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/27/2022	Water	N2511-06	Flash Point	Cool 4 deg C	PSEG03	P11	1530	04/20/2022	1010B
04/27/2022	Water	N2511-07	Flash Point	Cool 4 deg C	PSEG03	P11	1502	04/20/2022	1010B
04/27/2022	Water	N2511-08	Flash Point	Cool 4 deg C	PSEG03	P11	1551	04/20/2022	1010B
04/27/2022	Water	N2511-09	Flash Point	Cool 4 deg C	PSEG03	P11	1590	04/20/2022	1010B
04/27/2022	Water	N2511-10	Flash Point	Cool 4 deg C	PSEG03	P11	1593	04/20/2022	1010B
04/27/2022	Water	N2511-11	Flash Point	Cool 4 deg C	PSEG03	P11	1592	04/20/2022	1010B
04/27/2022	Water	N2511-12	Flash Point	Cool 4 deg C	PSEG03	P11	1591	04/20/2022	1010B
04/27/2022	Water	N2511-13	Flash Point	Cool 4 deg C	PSEG03	P11	1553	04/20/2022	1010B

Date/Time 04/21/22 14:45
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJC (EXT-1546)

Date/Time 04/21/22 18:00
 Raw Sample Received by: RJC (EXT-1546)
 Raw Sample Relinquished by: KS (WC)