



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119790

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	04/21/2022	17:37
2	CAL2	1	Water	NA	NA	20.4	7.00	04/21/2022	17:38
3	CAL3	1	Water	NA	NA	20.4	10.03	04/21/2022	17:39
4	ICV	1	Water	NA	NA	20.6	7.00	04/21/2022	17:40
5	CCV1	1	Water	NA	NA	20.6	2.01	04/21/2022	17:43
6	N2494-02	1	Solid	20.03	20	20.7	5.56	04/21/2022	17:45
7	N2494-05	1	Solid	20.04	20	20.3	5.58	04/21/2022	17:46
8	N2514-04	1	Solid	20.03	20	21.3	8.74	04/21/2022	17:48
9	N2514-08	1	Solid	20.04	20	21.4	9.62	04/21/2022	17:49
10	N2526-02	1	Solid	20.03	20	21.4	9.13	04/21/2022	17:50
11	N2526-05	1	Solid	20.02	20	21.1	7.14	04/21/2022	17:52
12	N2526-08	1	Solid	20.03	20	20.9	6.90	04/21/2022	17:55
13	N2526-08DUP	1	Solid	20.04	20	20.8	6.91	04/21/2022	17:56
14	CCV2	1	Water	NA	NA	20.8	12.03	04/21/2022	17:57

WB 119790

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2514 WorkList ID : 158873 Department : Wet-Chemistry Date : 04-21-2022 11:21:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
04/19/2022	Solid	N2494-02	Corrosivity	Cool 4 deg C	PSEG03	P11	MH-1S	04/19/2022	9045D
04/19/2022	Solid	N2494-05	Corrosivity	Cool 4 deg C	PSEG03	P11	MH-1D	04/19/2022	9045D
04/20/2022	Solid	N2514-04	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-16-1A-5	04/20/2022	9045D
04/20/2022	Solid	N2514-08	Corrosivity	Cool 4 deg C	ROUX02	P11	WC-16-A1-6	04/20/2022	9045D
04/20/2022	Solid	N2526-02	Corrosivity	Cool 4 deg C	PSEG03	P11	MH-3S	04/20/2022	9045D
04/21/2022	Solid	N2526-05	Corrosivity	Cool 4 deg C	PSEG03	P11	MH-5S	04/21/2022	9045D
04/21/2022	Solid	N2526-08	Corrosivity	Cool 4 deg C	PSEG03	P11	MH-5D	04/21/2022	9045D

Date/Time 04/21/22 16:50
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 04/21/22 18:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)