



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119793

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/21/2022	17:35
2	CAL2	1	Water	NA	NA	20.4	7.00	04/21/2022	17:36
3	CAL3	1	Water	NA	NA	20.4	10.03	04/21/2022	17:37
4	ICV	1	Water	NA	NA	20.5	6.99	04/21/2022	17:38
5	CCV1	1	Water	NA	NA	20.5	2.01	04/21/2022	17:39
6	N2519-01	1	Water	NA	NA	20.3	5.94	04/21/2022	17:42
7	N2519-01DUP	1	Water	NA	NA	20.4	5.95	04/21/2022	17:43
8	N2519-03	1	Water	NA	NA	20.4	6.65	04/21/2022	17:45
9	N2519-05	1	Water	NA	NA	20.3	6.72	04/21/2022	17:48
10	N2519-07	1	Water	NA	NA	20.5	6.59	04/21/2022	17:49
11	CCV2	1	Water	NA	NA	20.6	12.03	04/21/2022	17:50

VB 119793

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2519 WorkList ID : 158904 Department : Wet-Chemistry Date : 04-21-2021 17:30:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2022	Water	N2519-01	pH	Cool 4 deg C	LOCK01	P11	MW-1	04/20/2022	SM 4500-H B
04/20/2022	Water	N2519-03	pH	Cool 4 deg C	LOCK01	P11	MW-2	04/20/2022	SM 4500-H B
04/20/2022	Water	N2519-05	pH	Cool 4 deg C	LOCK01	P11	MW-3	04/20/2022	SM 4500-H B
04/20/2022	Water	N2519-07	pH	Cool 4 deg C	LOCK01	P11	MW-4	04/20/2022	SM 4500-H B

Date/Time 04/21/22 17:37
Raw Sample Received by: OP SM
Raw Sample Relinquished by: OP SM

Date/Time 04/21/22 18:25
Raw Sample Received by: OP (SM)
Raw Sample Relinquished by: OP (WC)