



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB119796

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP96409
Turbidity Calibration std, 20NTU	WP96405
Turbidity Calibration std, 40NTU	WP96404
Turbidity Calibration std, 1NTU	WP96408
Turbidity Calibration std, 5NTU	WP96407
Turbidity Calibration - CCV std, 10 NTU	WP96406
10 NTU Standard 500 ml	W2752

Intercept: 0.2731

Slope: 0.9665

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.155			04/22/2022	09:05
2	CAL2	1	1	1.150			04/22/2022	09:08
3	CAL3	5	1	5.090			04/22/2022	09:12
4	CAL4	10	1	10.100			04/22/2022	09:15
5	CAL5	20	1	19.800			04/22/2022	09:19
6	CAL6	40	1	38.800			04/22/2022	09:22



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB119796

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.950	10.012	04/22/2022	09:26
2	ICB		1	0.161	-0.116	04/22/2022	09:29
3	CCV1	10	1	10.200	10.271	04/22/2022	09:33
4	CCB1		1	0.162	-0.115	04/22/2022	09:37
5	LB119796BL		1	0.166	-0.111	04/22/2022	09:40
6	N2519-01		1	0.349	0.079	04/22/2022	09:43
7	N2519-03		1	6.890	6.846	04/22/2022	09:47
8	N2519-05		1	0.220	-0.055	04/22/2022	09:51
9	N2519-07		1	0.250	-0.024	04/22/2022	09:54
10	N2519-07DUP		1	0.248	-0.026	04/22/2022	09:58
11	CCV2	10	1	10.100	10.168	04/22/2022	10:02
12	CCB2		1	0.167	-0.110	04/22/2022	10:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-042222 WorkList ID : 158919 Department : Wet-Chemistry Date : 04-22-2022 09:30:59

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
04/22/2022	Water	N2519-01	Turbidity	Cool 4 deg C	LOCK01	P11	MW-1	04/20/2022	SM2130 B
04/22/2022	Water	N2519-03	Turbidity	Cool 4 deg C	LOCK01	P11	MW-2	04/20/2022	SM2130 B
04/22/2022	Water	N2519-05	Turbidity	Cool 4 deg C	LOCK01	P11	MW-3	04/20/2022	SM2130 B
04/22/2022	Water	N2519-07	Turbidity	Cool 4 deg C	LOCK01	P11	MW-4	04/20/2022	SM2130 B

Date/Time 4/22/22 9:31
 Raw Sample Received by: AP(WC)
 Raw Sample Relinquished by: AP(WC)

Date/Time 4/22/22 10:30
 Raw Sample Received by: AP(WC)
 Raw Sample Relinquished by: AP(WC)