



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB119797

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	04/22/2022	12:05
2	N2522-01		1	10	100.00	>212.0	>212.0	04/22/2022	12:25
3	N2522-01DUP		1	10	100.00	>212.0	>212.0	04/22/2022	12:45
4	N2522-02		1	12	100.00	>212.0	>212.0	04/22/2022	13:10
5	N2522-03		1	15	100.00	>212.0	>212.0	04/22/2022	13:35
6	N2524-01		1	12	100.00	>212.0	>212.0	04/22/2022	13:58
7	N2524-03		1	15	100.00	>212.0	>212.0	04/22/2022	14:20
8	N2524-05		1	12	100.00	>212.0	>212.0	04/22/2022	14:40
9	N2524-07		1	15	100.00	>212.0	>212.0	04/22/2022	15:05
10	N2524-09		1	14	100.00	>212.0	>212.0	04/22/2022	15:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB119797

WorkList Name : FP42222 WorkList ID : 158921 Department : Wet-Chemistry Date : 04-22-2022 11:35:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Customer Sample	Collect Date	Method
						Storage	Location			
04/28/2022	Water	N2522-01	Flash Point	Cool 4 deg C	PSEG03	P11	NWB-1756		04/21/2022	1010B
04/28/2022	Water	N2522-02	Flash Point	Cool 4 deg C	PSEG03	P11	NWB-1757		04/21/2022	1010B
04/28/2022	Water	N2522-03	Flash Point	Cool 4 deg C	PSEG03	P11	NWB-1766		04/21/2022	1010B
04/28/2022	Water	N2524-01	Flash Point	Cool 4 deg C	PSEG03	P11	33659		04/21/2022	1010B
04/28/2022	Water	N2524-03	Flash Point	Cool 4 deg C	PSEG03	P11	33655		04/21/2022	1010B
04/28/2022	Water	N2524-05	Flash Point	Cool 4 deg C	PSEG03	P11	33656-57		04/21/2022	1010B
04/28/2022	Water	N2524-07	Flash Point	Cool 4 deg C	PSEG03	P11	33658		04/21/2022	1010B
04/28/2022	Water	N2524-09	Flash Point	Cool 4 deg C	PSEG03	P11	33660		04/21/2022	1010B

Date/Time 04/22/22 11:40
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: KS(wc)

Date/Time 04/22/22 15:30
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: KS(wc)