



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 4/25/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:00  
In Date: 04/22/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:22  
Out Date: 04/23/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB119804

| Lab ID    | Client SampleID    | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments        |
|-----------|--------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|-----------------|
| N2528-01  | TP6-1.5-2.0        | 1      | 1.15            | 8.45          | 9.6                      | 8.9                        | 91.7    |                 |
| N2528-02  | TP6-3.5-4.0        | 2      | 1.11            | 8.77          | 9.88                     | 9.23                       | 92.6    |                 |
| N2528-03  | TP6-5.5-6.0        | 3      | 1.13            | 8.46          | 9.59                     | 9.09                       | 94.1    |                 |
| N2528-04  | TP6-7.5-8.0        | 4      | 1.15            | 8.60          | 9.75                     | 9.36                       | 95.5    |                 |
| N2528-05  | TP8-1.5-2.0        | 5      | 1.14            | 8.82          | 9.96                     | 9.16                       | 90.9    |                 |
| N2528-06  | TP8-3.5-4.0        | 6      | 1.15            | 8.51          | 9.66                     | 9.19                       | 94.5    |                 |
| N2528-07  | TP8-5.5-6.0        | 7      | 1.19            | 8.64          | 9.83                     | 9.42                       | 95.3    |                 |
| N2528-08  | TP8-7.5-8.0        | 8      | 1.17            | 8.81          | 9.98                     | 9.65                       | 96.3    |                 |
| N2528-09  | TP12-1.5-2.0       | 9      | 1.12            | 8.60          | 9.72                     | 8.88                       | 90.2    |                 |
| N2528-10  | TP12-3.5-4.0       | 10     | 1.15            | 8.79          | 9.94                     | 9.3                        | 92.7    |                 |
| N2528-11  | TP12-5.5-6.0       | 11     | 1.16            | 8.81          | 9.97                     | 9.47                       | 94.3    |                 |
| N2528-12  | TP12-7.5-8.0       | 12     | 1.11            | 8.75          | 9.86                     | 9.07                       | 91.0    |                 |
| N2528-12D | TP12-7.5-8.0DUP    | 13     | 1.14            | 8.74          | 9.88                     | 9.09                       | 91.0    |                 |
| N2532-01  | ELI-CON            | 14     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | CONCRETE SAMPLE |
| N2541-01  | CL-02-042222       | 15     | 1.13            | 8.80          | 9.93                     | 9.18                       | 91.5    |                 |
| N2541-02  | CL-02-042222-EPH-2 | 16     | 1.13            | 8.80          | 9.93                     | 9.18                       | 91.5    |                 |
| N2544-01  | S-1                | 17     | 1.14            | 8.72          | 9.86                     | 8.78                       | 87.6    |                 |
| N2544-02  | S-2                | 18     | 1.17            | 8.54          | 9.71                     | 8.63                       | 87.4    |                 |
| N2544-03  | S-3                | 19     | 1.11            | 8.63          | 9.74                     | 8.96                       | 91.0    |                 |
| N2544-04  | S-4                | 20     | 1.11            | 8.49          | 9.6                      | 8.92                       | 92.0    |                 |
| N2544-05  | S-5                | 21     | 1.18            | 8.54          | 9.72                     | 8.84                       | 89.7    |                 |
| N2544-06  | S-6                | 22     | 1.19            | 8.79          | 9.98                     | 8.8                        | 86.6    |                 |
| N2544-07  | S-7                | 23     | 1.11            | 8.52          | 9.63                     | 8.94                       | 91.9    |                 |
| N2545-01  | H-1                | 24     | 1.19            | 8.44          | 9.63                     | 8.47                       | 86.3    |                 |
| N2545-02  | H-2                | 25     | 1.17            | 8.72          | 9.89                     | 8.5                        | 84.1    |                 |
| N2545-03  | H-3                | 26     | 1.15            | 8.80          | 9.95                     | 8.65                       | 85.2    |                 |
| N2546-01  | S-5                | 27     | 1.13            | 8.70          | 9.83                     | 8.94                       | 89.8    |                 |
| N2546-02  | S-6                | 28     | 1.19            | 8.77          | 9.96                     | 8.74                       | 86.1    |                 |



# PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 4/25/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:00  
In Date: 04/22/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:22  
Out Date: 04/23/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB119804

| Lab ID   | Client SampleID       | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| N2547-05 | SVOC-GPC-BLANK        | 29     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-06 | PEST-GPC-BLANK        | 30     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-07 | PEST-GPC-BLANK-SPIKE  | 31     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-08 | PCB-GPC-BLANK         | 32     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-09 | PCB-GPC-BLANK-SPIKE   | 33     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-10 | SVOC-GPC2-BLANK       | 34     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-11 | PEST-GPC2-BLANK       | 35     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-12 | PEST-GPC2-BLANK-SPIKE | 36     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-13 | PCB-GCP2-BLANK        | 37     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2547-14 | PCB-GCP2-BLANK-SPIKE  | 38     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   |          |
| N2548-01 | MH-8S                 | 39     | 1.14            | 8.70          | 9.84                     | 9.37                       | 94.6    |          |
| N2548-03 | MH-8S-EPH             | 40     | 1.16            | 8.57          | 9.73                     | 9.24                       | 94.3    |          |
| N2548-04 | MH-8D                 | 41     | 1.13            | 8.46          | 9.59                     | 8.97                       | 92.7    |          |
| N2548-06 | MH-8D-EPH             | 42     | 1.16            | 8.48          | 9.64                     | 8.76                       | 89.6    |          |
| N2548-07 | TP-9                  | 43     | 1.14            | 8.81          | 9.95                     | 9.56                       | 95.6    |          |
| N2548-09 | TP-9-EPH              | 44     | 1.14            | 8.81          | 9.95                     | 9.56                       | 95.6    |          |
| N2548-10 | TP-11                 | 45     | 1.19            | 8.53          | 9.72                     | 8.37                       | 84.2    |          |
| N2548-12 | TP-11-EPH             | 46     | 1.19            | 8.53          | 9.72                     | 8.37                       | 84.2    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

## WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042222

WorkList ID : 158915

Department : Wet-Chemistry

Date : 04-22-2022 08:30:38

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample    | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|--------------------|--------------|--------------|
| 04/30/2022 | Solid  | N2528-01 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP6-1.5-2.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-02 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP6-3.5-4.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-03 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP6-5.5-6.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-04 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP6-7.5-8.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-05 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP8-1.5-2.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-06 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP8-3.5-4.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-07 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP8-5.5-6.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-08 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP8-7.5-8.0        | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-09 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP12-1.5-2.0       | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-10 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP12-3.5-4.0       | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-11 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP12-5.5-6.0       | 04/20/2022   | Chemtech -SO |
| 04/30/2022 | Solid  | N2528-12 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP12-7.5-8.0       | 04/20/2022   | Chemtech -SO |
| 04/27/2022 | Solid  | N2532-01 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | ELI-CON            | 04/22/2022   | Chemtech -SO |
| 04/27/2022 | Solid  | N2541-01 | Percent Solids | Cool 4 deg C | PSEG05   | P11                         | CL-02-042222       | 04/22/2022   | Chemtech -SO |
| 04/27/2022 | Solid  | N2541-02 | Percent Solids | Cool 4 deg C | PSEG05   | P11                         | CL-02-042222-EPH-2 | 04/22/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2544-01 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-1                | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2544-02 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-2                | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2544-03 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-3                | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2544-04 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-4                | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2544-05 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-5                | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2544-06 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-6                | 04/20/2022   | Chemtech -SO |

Date/Time 04/22/22 16:00

Date/Time 04/22/22

Raw Sample Received by: JPL/CL

Raw Sample Received by: JPL/CL

Raw Sample Relinquished by: JPL/CL

Raw Sample Relinquished by: JPL/CL



# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042222      WorkList ID : 158915      Department : Wet-Chemistry      Date : 04-22-2022 08:30:38

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample       | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|-----------------------|--------------|--------------|
| 04/29/2022 | Solid  | N2544-07 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-7                   | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2545-01 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | H-1                   | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2545-02 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | H-2                   | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2545-03 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | H-3                   | 04/20/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2546-01 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-5                   | 04/21/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2546-02 | Percent Solids | Cool 4 deg C | EARTH03  | P11                         | S-6                   | 04/21/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-05 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | SVOC-GPC-BLANK        | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-06 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PEST-GPC-BLANK        | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-07 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PEST-GPC-BLANK-SPIKE  | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-08 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PCB-GPC-BLANK         | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-09 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PCB-GPC-BLANK-SPIKE   | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-10 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | SVOC-GPC2-BLANK       | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-11 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PEST-GPC2-BLANK       | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-12 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PEST-GPC2-BLANK-SPIKE | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-13 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PCB-GCP2-BLANK        | 04/15/2022   | Chemtech -SO |
| 04/25/2022 | Solid  | N2547-14 | Percent Solids | Cool 4 deg C | CHEM02   | P11                         | PCB-GCP2-BLANK-SPIKE  | 04/15/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2548-01 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | MH-8S                 | 04/22/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2548-03 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | MH-8S-EPH             | 04/22/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2548-04 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | MH-8D                 | 04/22/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2548-06 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | MH-8D-EPH             | 04/22/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2548-07 | Percent Solids | Cool 4 deg C | PSEG03   | P11                         | TP-9                  | 04/22/2022   | Chemtech -SO |

Date/Time 04/22/22 17:20  
 Raw Sample Received by: J.D. CSM  
 Raw Sample Relinquished by: J.D. CSM

11/10/2024

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042222      WorkList ID : 158915      Department : Wet-Chemistry      Date : 04-22-2022 08:30:38

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample |           | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------|-----------|--------------|--------------|
|            |        |          |                |              |          | Storage    | Location  |              |              |
| 04/29/2022 | Solid  | N2548-09 | Percent Solids | Cool 4 deg C | PSEG03   | P-11       | TP-9-EPH  | 04/22/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2548-10 | Percent Solids | Cool 4 deg C | PSEG03   | P-11       | TP-11     | 04/22/2022   | Chemtech -SO |
| 04/29/2022 | Solid  | N2548-12 | Percent Solids | Cool 4 deg C | PSEG03   | P-11       | TP-11-EPH | 04/22/2022   | Chemtech -SO |

Date/Time 04/22/22 16:00  
Raw Sample Received by: JDC  
Raw Sample Relinquished by: JDC

Date/Time 04/22/22 17:20  
Raw Sample Received by: JDC  
Raw Sample Relinquished by: JDC