

LB119801


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56129



A= 0.0000e+000

B= 1.7618e-004

C= -3.2410e-002

Rho= 0.9997045

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.023	-0.023	55	0.000	55					-
0.2	0.200	0.167	-0.033	1132	0.0 %	1132					-17
2.5	2.500	2.486	-0.014	14292	0.0 %	14292					-1
5.0	5.000	5.069	0.069	28956	0.0 %	28956					1
7.5	7.500	7.645	0.145	43578	0.0 %	43578					2
10.0	10.000	9.856	-0.144	56129	0.0 %	56129					-1

LB119805 INSTRUMENT ID : CV1

Method: 7471B Operator: Admin Date of Analysis: 22 Apr 2022 14:13:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	55	-	0.0000	7471B	PPB	22 Apr 2022 14:17:46	S	Std
0.2 - 1	SO 2	1132	-	0.2000	7471B	PPB	22 Apr 2022 14:22:12	S	Std
2.5 - 1	SO 5	14292	-	2.5000	7471B	PPB	22 Apr 2022 14:24:29	S	Std
5.0 - 1	SO 5	28956	-	5.0000	7471B	PPB	22 Apr 2022 14:26:45	S	Std
7.5 - 1	SO 5	43578	-	7.5000	7471B	PPB	22 Apr 2022 14:29:01	S	Std
10.0 - 1	SO 0	56129	-	10.0000	7471B	PPB	22 Apr 2022 14:33:51	S	Std
ICV91 - 1	ICV91	21871	3.8208	-	7471B	PPB	22 Apr 2022 14:36:38	U	SMPL
ICB91 - 1	ICB91	-66	-0.0440	-	7471B	PPB	22 Apr 2022 14:38:52	U	SMPL
CCV22 - 1	CCV22	28888	5.0570	-	7471B	PPB	22 Apr 2022 14:41:07	U	SMPL
CCB22 - 1	CCB22	-66	-0.0440	-	7471B	PPB	22 Apr 2022 14:43:23	U	SMPL
CRA - 1	CRA	1146	0.1695	-	7471B	PPB	22 Apr 2022 14:45:38	U	SMPL
HighStd - 1	HighStd	56384	9.9011	-	7471B	PPB	22 Apr 2022 14:47:53	U	SMPL
ChkStd - 1	ChkStd	40254	7.0594	-	7471B	PPB	22 Apr 2022 14:50:09	U	SMPL
PB144311BL - 1	PBS	-132	-0.0557	-	7471B	PPB	22 Apr 2022 14:52:26	U	SMPL
PB144311BS - 1	LCSS	20205	3.5272	-	7471B	PPB	22 Apr 2022 14:54:43	U	SMPL
N2508-01 - 1	40576-SOL	1913	0.3046	-	7471B	PPB	22 Apr 2022 14:56:59	U	SMPL
N2511-05 - 1	1526-29	6107	1.0435	-	7471B	PPB	22 Apr 2022 14:59:16	U	SMPL
N2514-03 - 1	WC-16-1A-5	15757	2.7436	-	7471B	PPB	22 Apr 2022 15:01:34	U	SMPL
N2514-07 - 1	WC-16-A1-6	8772	1.5130	-	7471B	PPB	22 Apr 2022 15:03:52	U	SMPL
N2523-01 - 1	SU-01-042122	1067	0.1556	-	7471B	PPB	22 Apr 2022 15:06:07	U	SMPL
CCV23 - 1	CCV23	29272	5.1246	-	7471B	PPB	22 Apr 2022 15:08:23	U	SMPL
CCB23 - 1	CCB23	-14	-0.0349	-	7471B	PPB	22 Apr 2022 15:10:38	U	SMPL
N2523-03 - 1	SU-04-042122	7590	1.3048	-	7471B	PPB	22 Apr 2022 15:12:54	U	SMPL
N2525-01 - 1	HR-01-042122	4509	0.7620	-	7471B	PPB	22 Apr 2022 15:15:09	U	SMPL
N2525-03 - 1	HR-02-042122	6467	1.1069	-	7471B	PPB	22 Apr 2022 15:17:24	U	SMPL
N2525-05 - 1	HR-04-042122	2494	0.4070	-	7471B	PPB	22 Apr 2022 15:19:39	U	SMPL
N2526-01 - 1	MH-3S	1194	0.1779	-	7471B	PPB	22 Apr 2022 15:21:55	U	SMPL
N2526-04 - 1	MH-5S	1840	0.2918	-	7471B	PPB	22 Apr 2022 15:24:11	U	SMPL
N2526-07 - 1	MH-5D	487	0.0534	-	7471B	PPB	22 Apr 2022 15:26:28	U	SMPL
N2526-07DUP - 1	MH-5DDUP	580	0.0698	-	7471B	PPB	22 Apr 2022 15:28:45	U	SMPL
N2526-07MS - 1	MH-5DMS	23200	4.0549	-	7471B	PPB	22 Apr 2022 15:31:01	U	SMPL
N2526-07MSD - 1	MH-5DMSD	22971	4.0145	-	7471B	PPB	22 Apr 2022 15:33:17	U	SMPL
CCV24 - 1	CCV24	29214	5.1144	-	7471B	PPB	22 Apr 2022 15:35:32	U	SMPL
CCB24 - 1	CCB24	-11	-0.0343	-	7471B	PPB	22 Apr 2022 15:37:47	U	SMPL
N2526-07LX5 - 1		201	0.0030	-	7471B	PPB	22 Apr 2022 15:40:02	U	SMPL
N2526-07A - 1		23671	4.1379	-	7471B	PPB	22 Apr 2022 15:42:18	U	SMPL
CCV25 - 1	CCV25	29531	5.1703	-	7471B	PPB	22 Apr 2022 15:44:33	U	SMPL
CCB25 - 1	CCB25	-28	-0.0373	-	7471B	PPB	22 Apr 2022 15:46:49	U	SMPL