



### Analytical Summary Report

Analysis Method: 9045D  
Parameter: Corrosivity  
Run Number: LB119808  
BalanceID: WC SC-4

Analyst By : JIGNESH  
Supervisor Review By : apatel  
Slope : 99.1  
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/22/2022	16:30
2	CAL2	1	Water	NA	NA	20.3	7.00	04/22/2022	16:31
3	CAL3	1	Water	NA	NA	20.3	10.03	04/22/2022	16:33
4	ICV	1	Water	NA	NA	20.4	7.02	04/22/2022	16:35
5	CCV1	1	Water	NA	NA	20.6	2.01	04/22/2022	16:36
6	N2548-02	1	Solid	20.03	20	22.0	6.09	04/22/2022	16:39
7	N2548-02DUP	1	Solid	20.04	20	22.1	6.10	04/22/2022	16:40
8	N2548-05	1	Solid	20.02	20	21.8	7.09	04/22/2022	16:42
9	N2548-08	1	Solid	20.03	20	21.7	6.56	04/22/2022	16:45
10	N2548-11	1	Solid	20.05	20	21.3	7.22	04/22/2022	16:48
11	CCV2	1	Water	NA	NA	20.6	12.02	04/22/2022	16:50

# WORKLIST(Hardcopy Internal Chain)

119808

**WorkList Name :** CORROSIONITY N2548     
 **WorkList ID :** 158938     
 **Department :** Wet-Chemistry     
 **Date :** 04-22-2022 16:16:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/22/2022	Solid	N2548-02	Corrosivity	Cool 4 deg C	PSEG03	P11	MH-8S	04/22/2022	9045D
04/22/2022	Solid	N2548-05	Corrosivity	Cool 4 deg C	PSEG03	P11	MH-8D	04/22/2022	9045D
04/22/2022	Solid	N2548-08	Corrosivity	Cool 4 deg C	PSEG03	P11	TP-9	04/22/2022	9045D
04/22/2022	Solid	N2548-11	Corrosivity	Cool 4 deg C	PSEG03	P11	TP-11	04/22/2022	9045D

**Date/Time** 04/22/22 16:20  
**Raw Sample Received by:** JP (WC)  
**Raw Sample Relinquished by:** JD (SM)

**Date/Time** 04/22/22 17:50  
**Raw Sample Received by:** JD (SM)  
**Raw Sample Relinquished by:** JP (WC)