

LB119811


7470A

INSTRUMENT M: CVZ

Linear

 μ Abs.:

41572



A= 0.0000e+000

B= 2.2863e-004

C= 3.4613e-002

Rho= 0.9957928

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.040	0.040	22	0.000	22					28
0.2	0.200	0.256	0.056	969	0.0 %	969					-16
2.5	2.500	2.091	-0.409	8996	0.0 %	8996					7
5.0	5.000	5.348	0.348	23240	0.0 %	23240					6
7.5	7.500	7.925	0.425	34513	0.0 %	34513					-3
10.0	10.000	9.539	-0.461	41572	0.0 %	41572					

LB119811 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Apr 2022 10:25:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	22	-	0.0000	7470A	PPB	25 Apr 2022 10:38:08	S	Std
0.2 - 1	SD-2	969	-	0.2000	7470A	PPB	25 Apr 2022 10:42:35	S	Std
2.5 - 1	SD-5	8996	-	2.5000	7470A	PPB	25 Apr 2022 10:47:02	S	Std
5.0 - 1	SD-5	23240	-	5.0000	7470A	PPB	25 Apr 2022 10:49:18	S	Std
7.5 - 1	SD-5	34513	-	7.5000	7470A	PPB	25 Apr 2022 10:51:34	S	Std
10.0 - 1	SD-8	41572	-	10.0000	7470A	PPB	25 Apr 2022 10:53:50	S	Std
ICV92 - 1	ICV92	15766	3.6392	-7470A	PPB	PPB	25 Apr 2022 10:59:21	U	SMPL
ICB92 - 1	ICB92	-42	0.0250	-7470A	PPB	PPB	25 Apr 2022 11:01:36	U	SMPL
CCV26 - 1	CCV26	21892	5.0398	-7470A	PPB	PPB	25 Apr 2022 11:03:51	U	SMPL
CCB26 - 1	CCB26	-54	0.0223	-7470A	PPB	PPB	25 Apr 2022 11:06:06	U	SMPL
CRA - 1	CRA	808	0.2193	-7470A	PPB	PPB	25 Apr 2022 11:08:20	U	SMPL
HighStd - 1	HighStd	41449	9.5112	-7470A	PPB	PPB	25 Apr 2022 11:10:35	U	SMPL
ChkStd - 1	ChkStd	30578	7.0258	-7470A	PPB	PPB	25 Apr 2022 11:12:50	U	SMPL
PB144315BL - 1	PBW	-132	0.0044	-7470A	PPB	PPB	25 Apr 2022 11:15:07	U	SMPL
PB144315BS - 1	LCSW	16844	3.8857	-7470A	PPB	PPB	25 Apr 2022 11:17:23	U	SMPL
N2494-02 - 1	MH-1S	-44	0.0246	-7470A	PPB	PPB	25 Apr 2022 11:19:39	U	SMPL
N2494-05 - 1	MH-1D	9	0.0367	-7470A	PPB	PPB	25 Apr 2022 11:21:55	U	SMPL
N2514-04 - 1	WC-16-1A-5	4	0.0355	-7470A	PPB	PPB	25 Apr 2022 11:24:12	U	SMPL
N2514-08 - 1	WC-16-A1-6	133	0.0650	-7470A	PPB	PPB	25 Apr 2022 11:26:29	U	SMPL
N2526-02 - 1	MH-3S	19	0.0390	-7470A	PPB	PPB	25 Apr 2022 11:28:44	U	SMPL
CCV27 - 1	CCV27	21311	4.9070	-7470A	PPB	PPB	25 Apr 2022 11:30:59	U	SMPL
CCB27 - 1	CCB27	-62	0.0204	-7470A	PPB	PPB	25 Apr 2022 11:33:14	U	SMPL
N2526-05 - 1	MH-5S	34	0.0424	-7470A	PPB	PPB	25 Apr 2022 11:35:29	U	SMPL
N2526-08 - 1	MH-5D	25	0.0403	-7470A	PPB	PPB	25 Apr 2022 11:37:45	U	SMPL
N2526-08DUP - 1	MH-5DDUP	17	0.0385	-7470A	PPB	PPB	25 Apr 2022 11:39:59	U	SMPL
N2526-08MS - 1	MH-5DMS	14691	3.3935	-7470A	PPB	PPB	25 Apr 2022 11:42:15	U	SMPL
N2526-08MSD - 1	MH-5DMSD	15686	3.6209	-7470A	PPB	PPB	25 Apr 2022 11:44:31	U	SMPL
PB144272TB - 1	PB144272TB	-37	0.0262	-7470A	PPB	PPB	25 Apr 2022 11:46:47	U	SMPL
N2526-08LX5 - 1		-39	0.0257	-7470A	PPB	PPB	25 Apr 2022 11:49:03	U	SMPL
N2526-08A - 1		16706	3.8542	-7470A	PPB	PPB	25 Apr 2022 11:51:20	U	SMPL
CCV28 - 1	CCV28	21318	4.9086	-7470A	PPB	PPB	25 Apr 2022 11:53:37	U	SMPL
CCB28 - 1	CCB28	-73	0.0179	-7470A	PPB	PPB	25 Apr 2022 11:55:52	U	SMPL