

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 04/26/2022

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 04/25/2022 16:10 DATE & TIME OUT1: 04/26/2022 08:05

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 103 °C

DATE & TIME IN2: 04/26/2022 08:37 DATE & TIME OUT2: 04/26/2022 10:15

Run Number: LB119822

BalanceID: WCSC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB119822BL	LB119822BL	78.5634	78.5634	100	78.5635	78.5635	78.5635	0.0001	1
2	N2550-01	PUMP-STATION-207ST-TRAIN-FAC	81.4759	81.4759	100	81.6708	81.6708	81.6708	0.1949	1949
3	N2550-02	DYCKMAN-200ST-A51515T50SB	92.4919	92.4919	100	92.5986	92.5986	92.5986	0.1067	1067
4	N2550-02DUP	DYCKMAN-200ST-A51515T50SBDUP	96.9895	96.9895	100	97.1014	97.1014	97.1014	0.1119	1119

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB114822

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts n2550 WorkList ID : 158962 Department : Wet-Chemistry Date : 04-25-2022 10:26:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/30/2022	Water	N2550-01	TS <u>CS</u>	Cool 4 deg C	DIST01	P11	PUMP-STATION-207ST-TRAIN-FACILITY	04/23/2022	SM2540 B
04/30/2022	Water	N2550-02	TS <u>L</u>	Cool 4 deg C	DIST01	P11	DYCKMAN-200ST-A51515T50SB	04/23/2022	SM2540 B

Date/Time 04/25/22 13:00
 Raw Sample Received by: JP LWC
 Raw Sample Relinquished by: JP LWC

Date/Time 04/25/22 16:20
 Raw Sample Received by: JP LWC
 Raw Sample Relinquished by: JP LWC