



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119823

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/25/2022	15:00
2	CAL2	1	Water	NA	NA	20.3	7.00	04/25/2022	15:01
3	CAL3	1	Water	NA	NA	20.3	10.03	04/25/2022	15:02
4	ICV	1	Water	NA	NA	20.5	6.99	04/25/2022	15:09
5	CCV1	1	Water	NA	NA	20.5	2.01	04/25/2022	15:10
6	N2557-02	1	Water	NA	NA	20.6	7.63	04/25/2022	15:15
7	N2557-02DUP	1	Water	NA	NA	20.7	7.64	04/25/2022	15:16
8	CCV2	1	Water	NA	NA	20.7	12.03	04/25/2022	15:19

119823

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2557 WorkList ID : 158981 Department : Wet-Chemistry Date : 04-25-2022 14:43:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/25/2022	Water	N2557-02	Corrosivity	Cool 4 deg C	SEAC01	L11	PURGE-WATER	04/25/2022	9040C

Date/Time 04/25/22 14:50
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC

Date/Time 04/25/2022 16:10
Raw Sample Received by: JRS
Raw Sample Relinquished by: JRS