



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119826

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/25/2022	16:45
2	CAL2	1	Water	NA	NA	20.3	7.01	04/25/2022	16:46
3	CAL3	1	Water	NA	NA	20.4	10.03	04/25/2022	16:47
4	ICV	1	Water	NA	NA	20.5	6.98	04/25/2022	16:49
5	CCV1	1	Water	NA	NA	20.3	2.02	04/25/2022	17:05
6	N2552-03	1	Water	NA	NA	20.7	6.84	04/25/2022	17:15
7	N2552-05	1	Water	NA	NA	20.4	5.65	04/25/2022	17:18
8	N2558-01	1	Water	NA	NA	20.6	7.63	04/25/2022	17:20
9	N2558-01DUP	1	Water	NA	NA	20.7	7.64	04/25/2022	17:21
10	CCV2	1	Water	NA	NA	20.7	12.02	04/25/2022	17:24

WORKLIST(Hardcopy Internal Chain)

18-14826

WorkList Name : PH N2552 WorkList ID : 158986 Department : Wet-Chemistry Date : 04-25-2022 16:26:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/25/2022	Water	N2552-03	pH	Cool 4 deg C	PSEG03	K11	33640	04/25/2022	9040C
04/25/2022	Water	N2552-05	pH	Cool 4 deg C	PSEG03	K11	33645	04/25/2022	9040C
04/25/2022	Water	N2558-01	pH	Cool 4 deg C	PSEG03	K11	31699	04/25/2022	9040C

Date/Time 04/25/22 16:30
 Raw Sample Received by: RJ C (A/C)
 Raw Sample Relinquished by: RJ C (A/C)

Date/Time 04/25/22 18:00
 Raw Sample Received by: RJ C (A/C)
 Raw Sample Relinquished by: RJ C (A/C)