



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119827

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/25/2022	16:45
2	CAL2	1	Water	NA	NA	20.3	7.01	04/25/2022	16:46
3	CAL3	1	Water	NA	NA	20.4	10.03	04/25/2022	16:47
4	ICV	1	Water	NA	NA	20.5	6.98	04/25/2022	16:49
5	CCV1	1	Water	NA	NA	20.3	2.02	04/25/2022	17:05
6	N2552-01	1	Solid	20.02	20	20.4	5.56	04/25/2022	17:06
7	N2552-01DUP	1	Solid	20.03	20	20.5	5.57	04/25/2022	17:07
8	CCV2	1	Water	NA	NA	20.2	12.03	04/25/2022	17:10

VB 119424

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S N2552 WorkList ID : 158987 Department : Wet-Chemistry Date : 04-25-2022 16:27:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/25/2022	Solid	N2552-01	pH	Cool 4 deg C	PSEG03	K11	33641	04/25/2022	9045D

Date/Time 04/25/22 16:20
Raw Sample Received by: JP (WCC)
Raw Sample Relinquished by: RJ CETA-lab

Date/Time 04/25/22 18:00
Raw Sample Received by: RJ CETA-lab
Raw Sample Relinquished by: JP (WCC)